

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134828

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.02	02/27/2025	09:30
2	CAL2	1	Water	NA	NA	20.3	7.01	02/27/2025	09:31
3	CAL3	1	Water	NA	NA	20.1	10.02	02/27/2025	09:33
4	ICV	1	Water	NA	NA	20.3	7.00	02/27/2025	09:37
5	CCV1	1	Water	NA	NA	20.2	2.01	02/27/2025	09:40
6	Q1435-01	1	Water	NA	NA	20.2	7.41	02/27/2025	09:49
7	Q1435-01DUP	1	Water	NA	NA	20.3	7.42	02/27/2025	09:50
8	CCV2	1	Water	NA	NA	20.1	12.02	02/27/2025	10:00

WORKLIST(Hardcopy Internal Chain)

VB 134828

WorkList Name : ph w q1435

WorkList ID : 187912

Department : Wet-Chemistry

Date : 02-27-2025 08:11:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1435-01-65	286107	Water	pH	Cool 4 deg C	PSEG03	H21	02/26/2025	9040C

Date/Time 02/27/25 08:15
Raw Sample Received by: JB WEC
Raw Sample Relinquished by: CP SN

Date/Time 02/27/25 12:00
Raw Sample Received by: CP SN
Raw Sample Relinquished by: JB WEC