

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134843

Slope : 98.4

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/28/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	02/28/2025	08:11
3	CAL3	1	Water	NA	NA	20.2	10.02	02/28/2025	08:15
4	ICV	1	Water	NA	NA	20.3	7.02	02/28/2025	08:20
5	CCV1	1	Water	NA	NA	20.2	2.01	02/28/2025	09:22
6	Q1460-01	1	Solid	20.02	20	24.6	5.52	02/28/2025	09:30
7	Q1460-01DUP	1	Solid	20.03	20	24.7	5.53	02/28/2025	09:31
8	CCV2	1	Water	NA	NA	20.2	12.02	02/28/2025	09:35

WORKLIST(Hardcopy Internal Chain)

UP 134843

WorkList Name : ph q1464

WorkList ID : 187948

Department : Wet-Chemistry

Date : 02-28-2025 07:39:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1460-01 - A	60270	Solid	pH	Cool 4 deg C	PSEG03	H31	02/27/2025	9045D

02/28/25

02/28/25 07:55

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

JP 02/28/25

Date/Time 02/28/25 12:30

Raw Sample Received by:

Raw Sample Relinquished by:

JP 02/28/25