

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB134849

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	02/28/2025	10:00
2	Q1425-03		1	14	100.00	>212.0	>212.0	02/28/2025	10:30
3	Q1425-03DUP		1	14	100.00	>212.0	>212.0	02/28/2025	11:00
4	Q1425-06		1	15	100.00	>212.0	>212.0	02/28/2025	11:30
5	Q1425-08		1	15	100.00	>212.0	>212.0	02/28/2025	12:00
6	Q1425-11		1	14	100.00	>212.0	>212.0	02/28/2025	12:30
7	Q1425-15		1	13	100.00	>212.0	>212.0	02/28/2025	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26134849

WorkList Name : FP-1425

WorkList ID : 187898

Department : Wet-Chemistry

Date : 02-26-2025 11:46:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1425-03 A	0217-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	H31	02/25/2025	1010B
Q1425-06 A	0212-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	H31	02/25/2025	1010B
Q1425-08 C	0214	Water	Flash Point	Cool 4 deg C	PSEG03	H31	02/25/2025	1010B
Q1425-11 A	0219-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	H31	02/25/2025	1010B
Q1425-15 A	0224-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	H31	02/25/2025	1010B

Date/Time 02/28/2025 09:05
 Raw Sample Received by: 12(50)
 Raw Sample Relinquished by: JF(609)

Date/Time 02/28/2025 13:40
 Raw Sample Received by: JF(609)
 Raw Sample Relinquished by: 12(50)