


7470A

LB 134872

/ INSTRUMENT 113: CV4

 μ Abs.:

41351



Linear

A= 0.0000e+000

B= 2.4334e-004

C= -2.2747e-002

Rho= 0.9998527

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	-0.019	-0.019	16	0.000	16					
0.2	0.200	0.194	-0.006	890	0.0 %	890	0				
2.5	2.500	2.519	0.019	10445	0.0 %	10445					
5.0	5.000	5.086	0.086	20994	0.0 %	20994					
7.5	7.500	7.380	-0.120	30422	0.0 %	30422					
10.0	10.000	10.040	0.040	41351	0.0 %	41351					

41351
16
890
10445
20994
30422
41351

Method: 7470A

Operator: Admin

LB134872 INSTRUMENR ID : CV1

Date of Analysis: 03 Mar 2025 13:23:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	16	-	0.0000	7470A	PPB			
0.2 - 1	10.2	890	-	0.2000	7470A	PPB	03 Mar 2025 13:28:51	S	Std
2.5 - 1	52.5	10445	-	2.5000	7470A	PPB	03 Mar 2025 13:31:08	S	Std
5.0 - 1	55	20994	-	5.0000	7470A	PPB	03 Mar 2025 13:33:25	S	Std
7.5 - 1	57.5	30422	-	7.5000	7470A	PPB	03 Mar 2025 13:35:43	S	Std
10.0 - 1	510	41351	-	10.0000	7470A	PPB	03 Mar 2025 13:38:00	S	Std
ICV26 - 1	ICV26	17018	4.1185	-	7470A	PPB	03 Mar 2025 13:40:17	S	Std
ICB26 - 1	ICB26	-220	-0.0763	-	7470A	PPB	03 Mar 2025 13:43:02	U	SMPL
CCV51 - 1	CCV51	21362	5.1756	-	7470A	PPB	03 Mar 2025 13:45:19	U	SMPL
CCB51 - 1	CCB51	6	-0.0213	-	7470A	PPB	03 Mar 2025 13:47:37	U	SMPL
CRA - 1	CRA	939	0.2058	-	7470A	PPB	03 Mar 2025 13:53:40	U	SMPL
HighStd - 1	HighStd	41421	10.0568	-	7470A	PPB	03 Mar 2025 13:55:57	U	SMPL
ChkStd - 1	ChkStd	28177	6.8340	-	7470A	PPB	03 Mar 2025 13:58:12	U	SMPL
PB166958BL - 1	PBW	9	-0.0206	-	7470A	PPB	03 Mar 2025 14:00:28	U	SMPL
PB166958BS - 1	LCSW	15250	3.6882	-	7470A	PPB	03 Mar 2025 14:05:27	U	SMPL
Q1420-04 - 1	TP-1-WC	-242	-0.0816	-	7470A	PPB	03 Mar 2025 14:10:44	U	SMPL
Q1420-08 - 1	TP-2-WC	-32	-0.0305	-	7470A	PPB	03 Mar 2025 14:13:00	U	SMPL
Q1420-08DUP - 1	TP-2-WCDUP	-2	-0.0232	-	7470A	PPB	03 Mar 2025 14:15:18	U	SMPL
CCV52 - 1	CCV52	21394	5.1834	-	7470A	PPB	03 Mar 2025 14:17:35	U	SMPL
CCB52 - 1	CCB52	-235	-0.0799	-	7470A	PPB	03 Mar 2025 14:24:22	U	SMPL
Q1420-08MS - 1		16067	3.8871	-	7470A	PPB	03 Mar 2025 14:26:37	U	SMPL
Q1420-08MSD - 1		15546	3.7603	-	7470A	PPB	03 Mar 2025 14:31:26	U	SMPL
Q1420-08LX5 - 1		-189	-0.0687	-	7470A	PPB	03 Mar 2025 14:33:41	U	SMPL
Q1420-08A - 1		14425	3.4875	-	7470A	PPB	03 Mar 2025 14:35:59	U	SMPL
PB166879TB - 1	PB166879TB	-165	-0.0629	-	7470A	PPB	03 Mar 2025 14:38:16	U	SMPL
CCV53 - 1	CCV53	21259	5.1505	-	7470A	PPB	03 Mar 2025 14:40:32	U	SMPL
CCB53 - 1	CCB53	-213	-0.0746	-	7470A	PPB	03 Mar 2025 14:42:51	U	SMPL
				-	7470A	PPB	03 Mar 2025 14:45:07	U	SMPL