

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB134873

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/03/2025	13:40
2	CAL2	1	Water	NA	NA	20.3	7.01	03/03/2025	13:41
3	CAL3	1	Water	NA	NA	20.3	10.02	03/03/2025	13:45
4	ICV	1	Water	NA	NA	20.2	7.00	03/03/2025	13:48
5	CCV1	1	Water	NA	NA	20.2	2.01	03/03/2025	13:50
6	Q1467-01	1	Water	NA	NA	20.1	6.91	03/03/2025	14:00
7	Q1467-02	1	Water	NA	NA	20.1	7.74	03/03/2025	14:10
8	Q1467-03	1	Water	NA	NA	20.1	10.27	03/03/2025	14:20
9	Q1467-04	1	Water	NA	NA	20.2	9.51	03/03/2025	14:29
10	Q1470-02	1	Water	NA	NA	20.6	7.80	03/03/2025	14:37
11	Q1473-01	1	Water	NA	NA	20.7	7.84	03/03/2025	14:45
12	Q1473-02	1	Water	NA	NA	20.7	8.43	03/03/2025	14:50
13	Q1476-01	1	Water	NA	NA	20.4	8.30	03/03/2025	15:00
14	Q1476-04	1	Water	NA	NA	20.6	3.29	03/03/2025	15:10
15	Q1476-05	1	Water	NA	NA	20.2	5.27	03/03/2025	15:15
16	CCV2	1	Water	NA	NA	20.3	12.02	03/03/2025	15:20
17	Q1478-01	1	Water	NA	NA	20.3	6.68	03/03/2025	15:27
18	Q1478-03	1	Water	NA	NA	20.2	6.83	03/03/2025	15:33
19	Q1478-05	1	Water	NA	NA	20.2	6.59	03/03/2025	15:40
20	Q1478-07	1	Water	NA	NA	20.1	6.76	03/03/2025	15:55

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	Q1478-07DUP	1	Water	NA	NA	20.2	6.77	03/03/2025	15:56
22	CCV3	1	Water	NA	NA	20.2	2.01	03/03/2025	16:00

WORKLIST(Hardcopy Internal Chain)

134873

WorkList Name : PHN Q1478

WorkList ID : 187983

Department : Wet-Chemistry

Date : 03-03-2025 13:28:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1467-01	0227	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1467-02	50607	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1467-03	0226-A	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1467-04	0226-B	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1470-02	12925-DRIP	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1473-01	BUR-25-0011	Water	pH	Cool 4 deg C	PSEG03	J31	02/28/2025	9040C
Q1473-02	BUR-1293	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1476-01	TRE-25-0014	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1476-04	TRE-25-0023	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1476-05	TRE-25-0024	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1478-01	IDW-AQ-MW-19B-COMP-0228;	Water	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9040C
Q1478-03	IDW-AQ-IW-01-COMP-022825	Water	pH	Cool 4 deg C	JACO05	H31	02/28/2025	9040C
Q1478-05	IDW-AQ-IW-02-COMP-022825	Water	pH	Cool 4 deg C	JACO05	H31	02/28/2025	9040C
Q1478-07	IDW-AQ-IW-03-COMP-022825	Water	pH	Cool 4 deg C	JACO05	H31	02/28/2025	9040C

Date/Time 03/03/25 13:35

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 03/03/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]