

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134874

Slope : 98.8

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/03/2025	13:40
2	CAL2	1	Water	NA	NA	20.3	7.01	03/03/2025	13:41
3	CAL3	1	Water	NA	NA	20.3	10.02	03/03/2025	13:45
4	ICV	1	Water	NA	NA	20.2	7.00	03/03/2025	13:50
5	CCV1	1	Water	NA	NA	20.2	2.01	03/03/2025	16:10
6	Q1471-01	1	Solid	20.02	20	23.2	5.02	03/03/2025	16:20
7	Q1474-01	1	Solid	20.01	20	21.1	8.61	03/03/2025	16:30
8	Q1474-01DUP	1	Solid	20.02	20	21.2	8.62	03/03/2025	16:31
9	Q1476-02	1	Solid	20.03	20	20.5	4.97	03/03/2025	16:45
10	Q1476-03	1	Solid	20.01	20	21.0	6.17	03/03/2025	16:55
11	CCV2	1	Water	NA	NA	20.3	12.02	03/03/2025	17:00

WORKLIST(Hardcopy Internal Chain)

U134874

WorkList Name : PH S Q1478

WorkList ID : 187984

Department : Wet-Chemistry

Date : 03-03-2025 13:29:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1471-01	C 22225	Solid	pH	Cool 4 deg C	PSEG03	J31	02/28/2025	9045D
Q1474-01	F BU-03-02282025	Solid	pH	Cool 4 deg C	PSEG05	H31	02/28/2025	9045D
Q1476-02	D TRE-25-0016	Solid	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9045D
Q1476-03	D TRE-25-0017	Solid	pH	Cool 4 deg C	PSEG03	H31	02/28/2025	9045D

Date/Time 03/03/25 13:35

Raw Sample Received by: HWC

Raw Sample Relinquished by: J.C (SUN)

Date/Time 03/03/25

Raw Sample Received by: J.C (SUN)

Raw Sample Relinquished by: J.C (SUN)