

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134875

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	03/03/2025	11:15
2	Q1470-02		1	15	100.00	>212.0	>212.0	03/03/2025	11:45
3	Q1470-02DUP		1	15	100.00	>212.0	>212.0	03/03/2025	12:15
4	Q1470-03		1	5	7.00	44.6	43.9	03/03/2025	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134075

WorkList Name : fp-3-03

WorkList ID : 187987

Department : Wet-Chemistry

Date : 03-03-2025 09:46:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1470-02 D	12925-DRIP	Water	Flash Point	Cool 4 deg C	PSEG03	J31	02/28/2025	1010B
Q1470-03 B	1359-DIESEL	Water	Flash Point	Cool 4 deg C	PSEG03	J31	02/28/2025	1010B

Date/Time 03/03/2025 08:20
 Raw Sample Received by: RM WGS
 Raw Sample Relinquished by: DP WGS

Date/Time 03/03/2025 13:25
 Raw Sample Received by: DP WGS
 Raw Sample Relinquished by: RM WGS