

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134876

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	03/03/2025	08:35
2	Q1454-03		1	6	49.00	120.2	119.5	03/03/2025	09:05
3	Q1465-02		1	14	100.00	>212.0	>212.0	03/03/2025	09:35
4	Q1465-02DUP		1	14	100.00	>212.0	>212.0	03/03/2025	10:05
5	Q1470-01		1	5	7.00	44.6	43.9	03/03/2025	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-3-3

WorkList ID : 187981

Department : Wet-Chemistry

Date : 03-03-2025 08:11:13

15134876

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1454-03 A	LAW-25-0043	Solid	Flash Point	Cool 4 deg C	PSEG03	H31	02/27/2025	1010B
Q1465-02 B	W11-6327	Solid	Flash Point	Cool 4 deg C	PSEG03	H31	02/27/2025	1010B
Q1470-01 A	1359-GAS	Solid	Flash Point	Cool 4 deg C	PSEG03	J31	02/28/2025	1010B

Date/Time 03/03/2025 08.20
 Raw Sample Received by: R14 CWG
 Raw Sample Relinquished by: R14 CWG

Date/Time 03/03/2025 13.25
 Raw Sample Received by: R14 CWG
 Raw Sample Relinquished by: R14 CWG