


LB134897

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.029	0.029	39	0.000	39					-
0.05	0.050	0.064	0.014	188	0.0 %	188					28
0.2	0.200	0.234	0.034	905	0.0 %	905	0				17
2.5	2.500	2.400	-0.100	10089	0.0 %	10089					-4
5.0	5.000	4.998	-0.002	21105	0.0 %	21105					0
7.5	7.500	7.505	0.005	31734	0.0 %	31734					0
10.0	10.000	10.021	0.021	42401	0.0 %	42401					0

LB134877 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Mar 2025 15:00:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	245.1	PPB			
0.05 - 1	50.05	188	-	0.0500	245.1	PPB	03 Mar 2025 15:02:59	S	Std
0.2 - 1	50.2	905	-	0.2000	245.1	PPB	03 Mar 2025 15:05:16	S	Std
2.5 - 1	52.5	10089	-	2.5000	245.1	PPB	03 Mar 2025 15:07:33	S	Std
5.0 - 1	55	21105	-	5.0000	245.1	PPB	03 Mar 2025 15:09:50	S	Std
7.5 - 1	57.5	31734	-	7.5000	245.1	PPB	03 Mar 2025 15:12:07	S	Std
10.0 - 1	57	42401	-	10.0000	245.1	PPB	03 Mar 2025 15:14:23	S	Std
ICV27 - 1	ICV27	16662	3.9499	-245.1		PPB	03 Mar 2025 15:32:31	S	Std
ICB27 - 1	ICB27	-295	-0.0495	-245.1		PPB	03 Mar 2025 15:36:51	U	SMPL
CCV53 - 1	CCV53	21405	5.0686	-245.1		PPB	03 Mar 2025 15:39:06	U	SMPL
CCB53 - 1	CCB53	-319	-0.0552	-245.1		PPB	03 Mar 2025 15:41:23	U	SMPL
CRA - 1	CRA	120	0.0484	-245.1		PPB	03 Mar 2025 15:43:38	U	SMPL
HighStd - 1	HighStd	43098	10.1850	-245.1		PPB	03 Mar 2025 15:45:55	U	SMPL
ChkStd - 1	ChkStd	28706	6.7906	-245.1		PPB	03 Mar 2025 15:48:11	U	SMPL
PB166959BL - 1	PBW	-16	0.0163	-245.1		PPB	03 Mar 2025 15:50:26	U	SMPL
PB166959BS - 1	LCSW	15597	3.6987	-245.1		PPB	03 Mar 2025 15:55:22	U	SMPL
Q1439-01 - 1	LRSA-MOD	-226	-0.0332	-245.1		PPB	03 Mar 2025 15:57:37	U	SMPL
Q1439-01DUP - 1	LRSA-MODDUP	-59	0.0062	-245.1		PPB	03 Mar 2025 15:59:54	U	SMPL
Q1439-01MS - 1	LRSA-MODMS	18856	4.4674	-245.1		PPB	03 Mar 2025 16:02:12	U	SMPL
Q1439-01MSD - 1	LRSA-MODMSD	19039	4.5105	-245.1		PPB	03 Mar 2025 16:09:27	U	SMPL
CCV54 - 1	CCV54	20944	4.9598	-245.1		PPB	03 Mar 2025 16:11:44	U	SMPL
CCB54 - 1	CCB54	-323	-0.0561	-245.1		PPB	03 Mar 2025 16:14:01	U	SMPL
Q1439-01LX5 - 1		-36	0.0116	-245.1		PPB	03 Mar 2025 16:16:19	U	SMPL
Q1439-01A - 1		18017	4.2695	-245.1		PPB	03 Mar 2025 16:18:37	U	SMPL
CCV55 - 1	CCV55	21279	5.0389	-245.1		PPB	03 Mar 2025 16:20:52	U	SMPL
CCB55 - 1	CCB55	-354	-0.0634	-245.1		PPB	03 Mar 2025 16:23:07	U	SMPL
				-245.1		PPB	03 Mar 2025 16:25:25	U	SMPL