

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134888

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	03/04/2025	09:50
2	Q1467-01		1	15	100.00	>212.0	>212.0	03/04/2025	10:20
3	Q1467-01DUP		1	15	100.00	>212.0	>212.0	03/04/2025	10:50
4	Q1467-02		1	14	100.00	>212.0	>212.0	03/04/2025	11:20
5	Q1467-05		1	15	100.00	>212.0	>212.0	03/04/2025	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134880

WorkList Name : FP-03-03

WorkList ID : 187992

Department : Wet-Chemistry

Date : 03-03-2025 13:47:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1467-01 <u>B</u>	0227	Water	Flash Point	Cool 4 deg C	PSEG03	H31	02/28/2025	1010B
Q1467-02 <u>B</u>	50607	Water	Flash Point	Cool 4 deg C	PSEG03	H31	02/28/2025	1010B
Q1467-05 <u>B</u>	0226	Water	Flash Point	Cool 4 deg C	PSEG03	H31	02/28/2025	1010B

Date/Time 03/04/2025 09:40
 Raw Sample Received by: RMCCW
 Raw Sample Relinquished by: RMCCW

Date/Time 03/04/2025 12:30
 Raw Sample Received by: RMCCW
 Raw Sample Relinquished by: RMCCW