

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB134895

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/04/2025	16:35
2	CAL2	1	Water	NA	NA	20.2	7.00	03/04/2025	16:36
3	CAL3	1	Water	NA	NA	20.3	10.02	03/04/2025	16:39
4	ICV	1	Water	NA	NA	20.2	7.01	03/04/2025	16:44
5	CCV1	1	Water	NA	NA	20.3	2.01	03/04/2025	16:45
6	Q1480-33	1	Water	NA	NA	20.7	6.53	03/04/2025	16:55
7	Q1480-33DUP	1	Water	NA	NA	20.8	6.54	03/04/2025	16:56
8	CCV2	1	Water	NA	NA	20.3	12.02	03/04/2025	17:00

WORKLIST(Hardcopy Internal Chain)

VB 134895

WorkList Name : PH Q1480

WorkList ID : 188015

Department : Wet-Chemistry

Date : 03-04-2025 16:22:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1480-33	TP-3-WATER-SAMPLE	Water	pH	Cool 4 deg C	PSEG03	I11	03/03/2025	9040C

Date/Time 03/04/25 16:25

Raw Sample Received by: 40 [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: