

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134900

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/04/2025	15:00
2	CAL2	1	Water	NA	NA	20.3	7.00	03/04/2025	15:01
3	CAL3	1	Water	NA	NA	20.3	10.02	03/04/2025	15:05
4	ICV	1	Water	NA	NA	20.2	7.01	03/04/2025	15:10
5	CCV1	1	Water	NA	NA	20.2	2.02	03/04/2025	15:15
6	Q1478-13	1	Solid	20.02	20	20.9	7.10	03/04/2025	15:30
7	Q1481-02	1	Solid	20.03	20	21.2	7.17	03/04/2025	15:40
8	Q1484-06	1	Solid	20.02	20	22.3	7.28	03/04/2025	15:55
9	Q1484-06DUP	1	Solid	20.03	20	22.4	7.29	03/04/2025	15:56
10	CCV2	1	Water	NA	NA	20.3	12.02	03/04/2025	16:00

WORKLIST(Hardcopy Internal Chain)

UB 134900

WorkList Name : corrosion p1480

WorkList ID : 188003

Department : Wet-Chemistry

Date : 03-04-2025 11:38:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1478-13	IDW-SO-COMP-022825	Solid	Corrosivity	Cool 4 deg C	JACO05	H31	02/28/2025	9045D
Q1481-02	WC-K1311	Solid	Corrosivity	Cool 4 deg C	PSEG03	H31	03/04/2025	9045D
Q1484-06	TR-OTR-COMP-01	Solid	Corrosivity	Cool 4 deg C	PSEG03	H31	03/04/2025	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

03/04/25 14:30

SB WOC

AP Sm

Date/Time 03/04/25

Raw Sample Received by: AP Sm

Raw Sample Relinquished by: SB WOC

18:00

AP Sm

SB WOC