

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134902

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	03/05/2025	10:00
2	Q1476-01		1	14	100.00	>212.0	>212.0	03/05/2025	10:30
3	Q1476-01DUP		1	14	100.00	>212.0	>212.0	03/05/2025	11:00
4	Q1476-04		1	16	100.00	>212.0	>212.0	03/05/2025	11:30
5	Q1476-05		1	14	100.00	>212.0	>212.0	03/05/2025	12:00
6	Q1480-33		1	15	100.00	>212.0	>212.0	03/05/2025	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06