

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134922

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.3	03/06/2025	09:45
2	Q1465-03		1	5	22.00	71.6	72.5	03/06/2025	10:15
3	Q1473-01		1	15	100.00	>212.0	>212.0	03/06/2025	10:45
4	Q1473-02		1	14	100.00	>212.0	>212.0	03/06/2025	11:15
5	Q1494-01		1	16	100.00	>212.0	>212.0	03/06/2025	11:45
6	Q1494-01DUP		1	16	100.00	>212.0	>212.0	03/06/2025	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06