

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134935

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/07/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	6.99	03/07/2025	08:16
3	CAL3	1	Water	NA	NA	20.2	10.02	03/07/2025	08:19
4	ICV	1	Water	NA	NA	20.2	7.01	03/07/2025	08:22
5	CCV1	1	Water	NA	NA	20.2	2.01	03/07/2025	09:00
6	Q1515-01	1	Solid	20.02	20	23.4	8.32	03/07/2025	09:10
7	Q1515-01DUP	1	Solid	20.03	20	23.5	8.33	03/07/2025	09:11
8	Q1516-01	1	Solid	20.01	20	22.8	7.03	03/07/2025	09:18
9	Q1516-02	1	Solid	20.03	20	23.2	6.60	03/07/2025	09:22
10	Q1518-01	1	Solid	20.04	20	22.3	7.75	03/07/2025	09:30
11	CCV2	1	Water	NA	NA	20.2	12.02	03/07/2025	09:33

WORKLIST(Hardcopy Internal Chain)

VB 134935

WorkList Name : ph w q1515

WorkList ID : 188095

Department : Wet-Chemistry

Date : 03-07-2025 08:03:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1515-01	✓ AU-06-030625	Solid	pH	Cool 4 deg C	PSEG05	I21	03/06/2025	9045D
Q1516-01	✓ GAS-AUD-25-0004	Solid	pH	Cool 4 deg C	PSEG03	I21	03/06/2025	9045D
Q1516-02	✓ GAS-AUD-25-0025	Solid	pH	Cool 4 deg C	PSEG03	I21	03/06/2025	9045D
Q1518-01	✓ OILY-DEBRIS-COMP	Solid	pH	Cool 4 deg C	PSEG03	I21	03/06/2025	9045D

Date/Time 03/07/25 08:10

Raw Sample Received by: SB CWC

Raw Sample Relinquished by: CP SM

Date/Time 03/07/25

Raw Sample Received by: CP SM

Raw Sample Relinquished by: SB CWC