

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134957

Slope : 98.6

BalanceID:

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/10/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	03/10/2025	08:16
3	CAL3	1	Water	NA	NA	20.2	10.02	03/10/2025	08:20
4	ICV	1	Water	NA	NA	20.2	7.01	03/10/2025	08:25
5	CCV1	1	Water	NA	NA	20.1	2.02	03/10/2025	09:20
6	Q1523-02	1	Solid	20.02	20	23.9	12.30	03/10/2025	09:55
7	Q1523-05	1	Solid	20.03	20	24.0	12.27	03/10/2025	10:10
8	Q1523-05DUP	1	Solid	20.04	20	24.1	12.28	03/10/2025	10:11
9	Q1529-01	1	Solid	20.02	20	25.3	5.03	03/10/2025	10:25
10	Q1529-02	1	Solid	20.03	20	23.7	4.92	03/10/2025	10:30
11	Q1529-03	1	Solid	20.02	20	24.4	5.57	03/10/2025	10:35
12	Q1529-04	1	Solid	20.01	20	24.6	10.01	03/10/2025	10:48
13	CCV2	1	Water	NA	NA	20.3	12.02	03/10/2025	10:56

WORKLIST(Hardcopy Internal Chain)

493134957

WorkList Name : ph s q1529

WorkList ID : 188128

Department : Wet-Chemistry

Date : 03-10-2025 08:01:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1523-02	B WC-AI-01-C	Solid	pH	Cool 4 deg C	ENTA05	I31	03/06/2025	9045D
Q1523-05	B WC-AI-02-C	Solid	pH	Cool 4 deg C	ENTA05	I31	03/06/2025	9045D
Q1529-01	C 60300	Solid	pH	Cool 4 deg C	PSEG03	I31	03/07/2025	9045D
Q1529-02	A 60301	Solid	pH	Cool 4 deg C	PSEG03	I31	03/07/2025	9045D
Q1529-03	A 60303	Solid	pH	Cool 4 deg C	PSEG03	I31	03/07/2025	9045D
Q1529-04	D 60289	Solid	pH	Cool 4 deg C	PSEG03	I31	03/07/2025	9045D

Date/Time 03/10/25 08:15

Raw Sample Received by: 88 wci

Raw Sample Relinquished by: 88 wci

Date/Time 03/10/25

Raw Sample Received by: 88 wci

Raw Sample Relinquished by: 88 wci