

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134967

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/10/2025	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	03/10/2025	15:01
3	CAL3	1	Water	NA	NA	20.3	10.02	03/10/2025	15:05
4	ICV	1	Water	NA	NA	20.3	7.02	03/10/2025	15:15
5	CCV1	1	Water	NA	NA	20.2	2.01	03/10/2025	15:20
6	Q1536-01	1	Water	NA	NA	21.0	6.10	03/10/2025	15:28
7	Q1536-02	1	Water	NA	NA	21.3	6.17	03/10/2025	15:35
8	Q1536-03	1	Water	NA	NA	21.3	5.47	03/10/2025	15:44
9	Q1538-01	1	Water	NA	NA	20.7	7.67	03/10/2025	15:50
10	Q1538-01DUP	1	Water	NA	NA	20.8	7.68	03/10/2025	15:51
11	CCV2	1	Water	NA	NA	20.3	12.02	03/10/2025	16:00

WORKLIST(Hardcopy Internal Chain)

134967

WorkList Name : PH Q1538

WorkList ID : 188166

Department : Wet-Chemistry

Date : 03-10-2025 14:40:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1536-01	A 60425	Water	pH	Cool 4 deg C	PSEG03	I31	03/10/2025	9040C
Q1536-02	A 60426	Water	pH	Cool 4 deg C	PSEG03	I31	03/10/2025	9040C
Q1536-03	A 60427	Water	pH	Cool 4 deg C	PSEG03	I31	03/10/2025	9040C
Q1538-01	A BELL-25-0004	Water	pH	Cool 4 deg C	PSEG03	H21	03/10/2025	9040C

Date/Time 03/10/25 14:50

Raw Sample Received by: JB GRC

Raw Sample Relinquished by: JTCMY

Date/Time 03/10/25 17:30

Raw Sample Received by: JTCMY

Raw Sample Relinquished by: JB GRC