

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134968

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/10/2025	14:10
2	CAL2	1	Water	NA	NA	20.2	7.00	03/10/2025	14:11
3	CAL3	1	Water	NA	NA	20.3	10.02	03/10/2025	14:11
4	ICV	1	Water	NA	NA	20.3	7.01	03/10/2025	14:15
5	CCV1	1	Water	NA	NA	20.2	2.01	03/10/2025	14:20
6	Q1534-06	1	Solid	20.02	20	23.6	5.89	03/10/2025	14:29
7	Q1534-06DUP	1	Solid	20.03	20	23.8	5.90	03/10/2025	14:30
8	Q1534-12	1	Solid	20.02	20	23.6	6.29	03/10/2025	14:33
9	Q1534-18	1	Solid	20.04	20	22.8	6.53	03/10/2025	14:37
10	Q1534-24	1	Solid	20.02	20	22.6	6.68	03/10/2025	14:45
11	CCV2	1	Water	NA	NA	20.3	12.02	03/10/2025	14:50

WORKLIST(Hardcopy Internal Chain)

20124968

WorkList Name : corrosivity p1534 WorkList ID : 188164 Department : Wet-Chemistry Date : 03-10-2025 13:45:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1534-06	OR-363-COMP-16	Solid	Corrosivity	Cool 4 deg C	PSEG03	H31	03/07/2025	9045D
Q1534-12	OR-363-COMP-17	Solid	Corrosivity	Cool 4 deg C	PSEG03	H31	03/07/2025	9045D
Q1534-18	OR-363-COMP-18	Solid	Corrosivity	Cool 4 deg C	PSEG03	H31	03/07/2025	9045D
Q1534-24	OR-363OR-363-COMP-19-	Solid	Corrosivity	Cool 4 deg C	PSEG03	H31	03/07/2025	9045D

Date/Time 03/10/25 13:55

Raw Sample Received by: SP

Raw Sample Relinquished by: SP

Date/Time 03/10/25 17:13

Raw Sample Received by: SP

Raw Sample Relinquished by: SP