

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134973

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	03/12/2025	12:30
2	Q1510-02		1	5	7.00	44.6	44.6	03/12/2025	13:00
3	Q1510-02DUP		1	5	7.00	44.6	44.6	03/12/2025	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134973

WorkList Name : flash point-3-11

WorkList ID : 188173

Department : Wet-Chemistry

Date : 03-11-2025 08:02:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1510-02 B	FMC-25-0006	Solid	Flash Point	Cool 4 deg C	PSEG03	I21	03/06/2025	1010B

Date/Time 03/12/2025 12.20
Raw Sample Received by: RM (wcs)
Raw Sample Relinquished by: Muelc

Date/Time 03/12/2025 14.10
Raw Sample Received by: Muelc
Raw Sample Relinquished by: RM (wcs)