


LB134981

7470A INSTRUMENT ID: CV1

Linear

 μ Abs.:

56704



A= 0.0000e+000

B= 1.7596e-004

C= 1.7665e-002

Rho= 0.9998462

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.029	0.029	62	0.000	62	0				18
0.2	0.200	0.236	0.036	1238	0.0 %	1238					-1
2.5	2.500	2.466	-0.034	13916	0.0 %	13916					-2
5.0	5.000	4.883	-0.117	27652	0.0 %	27652					-1
7.5	7.500	7.591	0.091	43040	0.0 %	43040					0
10.0	10.000	9.995	-0.005	56704	0.0 %	56704					

LB134981 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Mar 2025 10:39:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	62	-	0.0000	7470A	PPB	11 Mar 2025 10:49:36	S	Std
0.2 - 1	20.2	1238	-	0.2000	7470A	PPB	11 Mar 2025 10:51:53	S	Std
2.5 - 1	52.5	13916	-	2.5000	7470A	PPB	11 Mar 2025 10:54:10	S	Std
5.0 - 1	55	27652	-	5.0000	7470A	PPB	11 Mar 2025 10:56:27	S	Std
7.5 - 1	57.5	43040	-	7.5000	7470A	PPB	11 Mar 2025 11:01:29	S	Std
10.0 - 1	570	56704	-	10.0000	7470A	PPB	11 Mar 2025 11:06:40	S	Std
ICV37 - 1	ICV37	22647	4.0026	-	7470A	PPB	11 Mar 2025 11:16:02	U	SMPL
ICB37 - 1	ICB37	-246	-0.0256	-	7470A	PPB	11 Mar 2025 11:18:16	U	SMPL
CCV84 - 1	CCV84	26838	4.7401	-	7470A	PPB	11 Mar 2025 11:20:34	U	SMPL
CCB84 - 1	CCB84	-211	-0.0195	-	7470A	PPB	11 Mar 2025 11:22:49	U	SMPL
CRA - 1	CRA	1169	0.2234	-	7470A	PPB	11 Mar 2025 11:25:07	U	SMPL
HighStd - 1	HighStd	54969	9.6900	-	7470A	PPB	11 Mar 2025 11:27:22	U	SMPL
ChkStd - 1	ChkStd	38827	6.8497	-	7470A	PPB	11 Mar 2025 11:29:37	U	SMPL
PB167082BL - 1	PBW	-459	-0.0631	-	7470A	PPB	11 Mar 2025 11:31:56	U	SMPL
PB167082BS - 1	LCSW	19683	3.4811	-	7470A	PPB	11 Mar 2025 11:34:15	U	SMPL
Q1528-02 - 1	22725	1432	0.2696	-	7470A	PPB	11 Mar 2025 11:36:31	U	SMPL
Q1538-01 - 1	BELL-25-0004	-167	-0.0117	-	7470A	PPB	11 Mar 2025 11:38:49	U	SMPL
Q1538-01DUP - 1	BELL-25-0004DUP	96	0.0346	-	7470A	PPB	11 Mar 2025 11:41:09	U	SMPL
Q1538-01MS - 1	BELL-25-0004MS	7292	1.3008	-	7470A	PPB	11 Mar 2025 11:43:26	U	SMPL
Q1538-01MSD - 1	BELL-25-0004MSD	6273	1.1215	-	7470A	PPB	11 Mar 2025 11:45:42	U	SMPL
CCV85 - 1	CCV85	26885	4.7484	-	7470A	PPB	11 Mar 2025 11:54:01	U	SMPL
CCB85 - 1	CCB85	-364	-0.0464	-	7470A	PPB	11 Mar 2025 11:56:16	U	SMPL
Q1538-01LX5 - 1		315	0.0731	-	7470A	PPB	11 Mar 2025 11:58:34	U	SMPL
Q1538-01A - 1		16110	2.8524	-	7470A	PPB	11 Mar 2025 12:00:50	U	SMPL
CCV86 - 1	CCV86	27246	4.8119	-	7470A	PPB	11 Mar 2025 12:06:21	U	SMPL
CCB86 - 1	CCB86	-395	-0.0518	-	7470A	PPB	11 Mar 2025 12:08:37	U	SMPL