

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134984

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/11/2025	12:30
2	CAL2	1	Water	NA	NA	20.2	7.00	03/11/2025	12:31
3	CAL3	1	Water	NA	NA	20.2	10.02	03/11/2025	12:33
4	ICV	1	Water	NA	NA	20.3	6.99	03/11/2025	12:37
5	CCV1	1	Water	NA	NA	20.2	2.01	03/11/2025	12:40
6	Q1505-20	1	Water	NA	NA	23.4	5.45	03/11/2025	12:50
7	Q1541-01	1	Water	NA	NA	20.3	6.35	03/11/2025	13:00
8	Q1541-01DUP	1	Water	NA	NA	20.4	6.36	03/11/2025	13:01
9	CCV2	1	Water	NA	NA	20.3	12.02	03/11/2025	13:10

WORKLIST(Hardcopy Internal Chain)

VB 13484

WorkList Name : PH W Q1541

WorkList ID : 188197

Department : Wet-Chemistry

Date : 03-11-2025 11:21:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1505-20	PT-PH-WP	Water	pH	Cool 4 deg C	ALLI03	QA 01	03/03/2025	9040C
Q1541-01 -D	205703	Water	pH	Cool 4 deg C	PSEG03	H31	03/11/2025	9040C

Date/Time 03/11/25 12:10

Raw Sample Received by: RM 501

Raw Sample Relinquished by: RM 501

Date/Time 03/11/25

Raw Sample Received by: RM 501

Raw Sample Relinquished by: RM 501