

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134989

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	03/11/2025	14:05
2	Q1528-01		1	7	72.00	161.6	161.9	03/11/2025	14:35
3	Q1528-02		1	14	100.00	>212.0	>212.0	03/11/2025	15:05
4	Q1528-03		1	16	100.00	>212.0	>212.0	03/11/2025	15:35
5	Q1541-01		1	15	100.00	>212.0	>212.0	03/11/2025	16:05
6	Q1541-01DUP		1	15	100.00	>212.0	>212.0	03/11/2025	16:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134989

WorkList Name : fp-3-11

WorkList ID : 188172

Department : Wet-Chemistry

Date : 03-11-2025 07:59:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1528-01 E	FERNOT	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/07/2025	1010B
Q1528-02 C	22725	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/07/2025	1010B
Q1528-03 E	22425	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/07/2025	1010B
Q1541-01 F	205703	Water	Flash Point	Cool 4 deg C	PSEG03	H31	03/11/2025	1010B

Date/Time 03/11/2025 13:40
 Raw Sample Received by: RM cwc
 Raw Sample Relinquished by: [Signature]

Date/Time 03/11/2025 13:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: RM cwc