

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134992

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/12/2025	08:30
2	CAL2	1	Water	NA	NA	20.3	7.00	03/12/2025	08:31
3	CAL3	1	Water	NA	NA	20.3	10.02	03/12/2025	08:35
4	ICV	1	Water	NA	NA	20.1	6.98	03/12/2025	08:44
5	CCV1	1	Water	NA	NA	20.3	2.01	03/12/2025	08:45
6	Q1545-01	1	Solid	20.02	20	22.9	8.25	03/12/2025	09:00
7	Q1545-01DUP	1	Solid	20.03	20	22.8	8.26	03/12/2025	09:01
8	CCV2	1	Water	NA	NA	20.3	12.02	03/12/2025	09:05

WORKLIST(Hardcopy Internal Chain)

134992

WorkList Name : PH S Q1545 WorkList ID : 188213 Department : Wet-Chemistry Date : 03-12-2025 08:18:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1545-01	BU-02-031125	Solid	pH	Cool 4 deg C	PSEG05	I41	03/11/2025	9045D

Date/Time 03/12/25 08:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 03/12/25 11:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]