

LA134996


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

47909



A= 0.0000e+000

B= 2.1091e-004

C= 4.3678e-002

Rho= 0.9993816

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/a
0.0	0.000	0.009	0.009	-164	0.000	-164	0				-
0.2	0.200	0.230	0.030	883	0.0 %	883					15
2.5	2.500	2.495	-0.005	11622	0.0 %	11622					0
5.0	5.000	5.084	0.084	23899	0.0 %	23899					2
7.5	7.500	7.234	-0.266	34090	0.0 %	34090					-4
10.0	10.000	10.148	0.148	47909	0.0 %	47909					1

LB134996 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Mar 2025 10:26:20

Sample ID	Extended ID	µ Abs	Conc	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	-164	-	0.00007470A	PPB	12 Mar 2025 10:41:11	S	Std
0.2 - 1	802	883	-	0.20007470A	PPB	12 Mar 2025 10:43:28	S	Std
2.5 - 1	72.5	11622	-	2.50007470A	PPB	12 Mar 2025 10:45:45	S	Std
5.0 - 1	55	23899	-	5.00007470A	PPB	12 Mar 2025 10:48:02	S	Std
7.5 - 1	325	34090	-	7.50007470A	PPB	12 Mar 2025 10:50:20	S	Std
10.0 - 1	56	47909	-	10.00007470A	PPB	12 Mar 2025 10:55:16	S	Std
ICV38 - 1	ICV38	17083	3.6467	-7470A	PPB	12 Mar 2025 11:00:40	U	SMPL
ICB38 - 1	ICB38	-284	-0.0162	-7470A	PPB	12 Mar 2025 11:02:56	U	SMPL
CCV86 - 1	CCV86	22775	4.8472	-7470A	PPB	12 Mar 2025 11:10:37	U	SMPL
CCB86 - 1	CCB86	-268	-0.0128	-7470A	PPB	12 Mar 2025 11:12:52	U	SMPL
CRA - 1	CRA	830	0.2187	-7470A	PPB	12 Mar 2025 11:15:09	U	SMPL
HighStd - 1	HighStd	44722	9.4760	-7470A	PPB	12 Mar 2025 11:17:24	U	SMPL
ChkStd - 1	ChkStd	30059	6.3835	-7470A	PPB	12 Mar 2025 11:25:36	U	SMPL
PB167093BL - 1	PBW	-135	0.0152	-7470A	PPB	12 Mar 2025 11:27:52	U	SMPL
PB167093BS - 1	LCSW	19950	4.2514	-7470A	PPB	12 Mar 2025 11:30:11	U	SMPL
Q1523-03 - 1	WC-A1-01-C	-218	-0.0023	-7470A	PPB	12 Mar 2025 11:32:27	U	SMPL
Q1523-06 - 1	WC-A2-02-C	-25	0.0384	-7470A	PPB	12 Mar 2025 11:34:47	U	SMPL
Q1534-06 - 1	OR-363-COMP-16	-23	0.0388	-7470A	PPB	12 Mar 2025 11:37:04	U	SMPL
Q1534-12 - 1	OR-363-COMP-17	-5	0.0426	-7470A	PPB	12 Mar 2025 11:39:21	U	SMPL
Q1534-18 - 1	OR-363-COMP-18	1	0.0439	-7470A	PPB	12 Mar 2025 11:41:36	U	SMPL
CCV87 - 1	CCV87	22568	4.8035	-7470A	PPB	12 Mar 2025 11:43:51	U	SMPL
CCB87 - 1	CCB87	-290	-0.0175	-7470A	PPB	12 Mar 2025 11:46:06	U	SMPL
Q1534-24 - 1	OR-363OR-363-COMP-19-	50	0.0542	-7470A	PPB	12 Mar 2025 11:48:24	U	SMPL
Q1537-05 - 1	COMP-1	30	0.0500	-7470A	PPB	12 Mar 2025 11:50:40	U	SMPL
Q1537-06 - 1	COMP-2	-1	0.0435	-7470A	PPB	12 Mar 2025 11:52:55	U	SMPL
Q1537-07 - 1	COMP-3	20	0.0479	-7470A	PPB	12 Mar 2025 11:55:10	U	SMPL
Q1537-07DUP - 1	COMP-3DUP	-27	0.0380	-7470A	PPB	12 Mar 2025 11:57:26	U	SMPL
Q1537-07MS - 1	COMP-3MS	14456	3.0926	-7470A	PPB	12 Mar 2025 11:59:43	U	SMPL
Q1537-07MSD - 1	COMP-3MSD	13776	2.9492	-7470A	PPB	12 Mar 2025 12:02:00	U	SMPL
PB167102BL - 1	PBW	103	0.0654	-7470A	PPB	12 Mar 2025 12:04:19	U	SMPL
Q1168-07 - 1	LOD-MDL-WATER-01-QT1-2025	19	0.0477	-7470A	PPB	12 Mar 2025 12:32:19	U	SMPL
CCV88 - 1	CCV88	22439	4.7763	-7470A	PPB	12 Mar 2025 12:54:14	U	SMPL
CCB88 - 1	CCB88	-143	0.0135	-7470A	PPB	12 Mar 2025 12:56:45	U	SMPL
Q1168-08 - 1	LOQ-WATER-02-QT1-2025	836	0.2200	-7470A	PPB	12 Mar 2025 13:01:51	U	SMPL
Q1539-01 - 1	TAPIAL3-MW03D-031025-00-T1	265	0.0996	-7470A	PPB	12 Mar 2025 13:04:07	U	SMPL
Q1539-02 - 1	TAPFTA-MW011-031025-00-T2	99	0.0646	-7470A	PPB	12 Mar 2025 13:06:22	U	SMPL
Q1541-01 - 1	205703	-131	0.0160	-7470A	PPB	12 Mar 2025 13:08:38	U	SMPL
Q1541-01DUP - 1	205703DUP	-39	0.0355	-7470A	PPB	12 Mar 2025 13:10:54	U	SMPL
Q1541-01MS - 1	205703MS	18459	3.9369	-7470A	PPB	12 Mar 2025 13:13:11	U	SMPL
Q1541-01MSD - 1	205703MSD	17915	3.8222	-7470A	PPB	12 Mar 2025 13:15:28	U	SMPL
CCV89 - 1	CCV89	22889	4.8712	-7470A	PPB	12 Mar 2025 13:17:46	U	SMPL
CCB89 - 1	CCB89	-135	0.0152	-7470A	PPB	12 Mar 2025 13:20:05	U	SMPL
PB167102BS - 1	LCSW	17016	3.6325	-7470A	PPB	12 Mar 2025 13:22:22	U	SMPL
PB167049TB - 1	PB167049TB	-249	-0.0088	-7470A	PPB	12 Mar 2025 13:24:38	U	SMPL
Q1537-07LX5 - 1		3	0.0443	-7470A	PPB	12 Mar 2025 13:26:56	U	SMPL
Q1537-07A - 1		18159	3.8736	-7470A	PPB	12 Mar 2025 13:29:12	U	SMPL
Q1541-01LX5 - 1		-101	0.0224	-7470A	PPB	12 Mar 2025 13:31:27	U	SMPL
Q1541-01A - 1		18521	3.9500	-7470A	PPB	12 Mar 2025 13:33:45	U	SMPL
CCV90 - 1	CCV90	24596	5.2313	-7470A	PPB	12 Mar 2025 13:46:04	U	SMPL
CCB90 - 1	CCB90	-165	0.0089	-7470A	PPB	12 Mar 2025 13:48:19	U	SMPL