

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB134999

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	03/12/2025	10:00
2	Q1478-01		1	13	100.00	>212.0	>212.0	03/12/2025	10:30
3	Q1478-01DUP		1	13	100.00	>212.0	>212.0	03/12/2025	11:00
4	Q1478-03		1	15	100.00	>212.0	>212.0	03/12/2025	11:30
5	Q1478-05		1	15	100.00	>212.0	>212.0	03/12/2025	12:00
6	Q1478-07		1	16	100.00	>212.0	>212.0	03/12/2025	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134999

WorkList Name : fp-3-12

WorkList ID : 188232

Department : Wet-Chemistry

Date : 03-12-2025 08:52:09

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1478-01	IDW-AQ-MW-19B-COMP-022825	Water	Flash Point	Cool 4 deg C	JACO05	H31	02/28/2025	1010B
Q1478-03	IDW-AQ-IW-01-COMP-022825	Water	Flash Point	Cool 4 deg C	JACO05	H31	02/28/2025	1010B
Q1478-05	IDW-AQ-IW-02-COMP-022825	Water	Flash Point	Cool 4 deg C	JACO05	H31	02/28/2025	1010B
Q1478-07	IDW-AQ-IW-03-COMP-022825	Water	Flash Point	Cool 4 deg C	JACO05	H31	02/28/2025	1010B

Date/Time 03/12/2025 09:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 03/12/2025 13:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]