

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135008

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/12/2025	13:00
2	CAL2	1	Water	NA	NA	20.3	7.00	03/12/2025	13:01
3	CAL3	1	Water	NA	NA	20.2	10.02	03/12/2025	13:02
4	ICV	1	Water	NA	NA	20.3	7.02	03/12/2025	13:05
5	CCV1	1	Water	NA	NA	20.3	2.01	03/12/2025	13:10
6	Q1547-05	1	Solid	20.02	20	23.4	6.31	03/12/2025	13:30
7	Q1547-05DUP	1	Solid	20.03	20	23.5	6.32	03/12/2025	13:31
8	Q1547-10	1	Solid	20.04	20	23.3	6.50	03/12/2025	13:40
9	Q1552-05	1	Solid	20.02	20	22.3	6.17	03/12/2025	13:45
10	CCV2	1	Water	NA	NA	20.2	12.02	03/12/2025	13:50

WORKLIST(Hardcopy Internal Chain)

135008

WorkList Name : corrosivity p1547 WorkList ID : 188229 Department : Wet-Chemistry Date : 03-12-2025 11:37:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1547-05	OR-620-JB-COMP-01	Solid	Corrosivity	Cool 4 deg C	PSEG03	I41	03/11/2025	9045D
Q1547-10	OR-620-JB-COMP-02	Solid	Corrosivity	Cool 4 deg C	PSEG03	I41	03/11/2025	9045D
Q1552-05	NB-418-JB-COMP	Solid	Corrosivity	Cool 4 deg C	PSEG03	I31	03/12/2025	9045D

Date/Time 03/12/25 12:30
 Raw Sample Received by: JP WOC
 Raw Sample Relinquished by: JP SM

Date/Time 03/12/25 16:00
 Raw Sample Received by: JP SM
 Raw Sample Relinquished by: JP WOC