

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135015

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/13/2025	08:50
2	CAL2	1	Water	NA	NA	20.3	7.00	03/13/2025	08:51
3	CAL3	1	Water	NA	NA	20.3	10.02	03/13/2025	08:52
4	ICV	1	Water	NA	NA	20.3	7.01	03/13/2025	09:00
5	CCV1	1	Water	NA	NA	20.2	2.02	03/13/2025	09:45
6	Q1553-03	1	Water	NA	NA	20.8	6.80	03/13/2025	09:55
7	Q1553-03DUP	1	Water	NA	NA	20.9	6.81	03/13/2025	09:56
8	CCV2	1	Water	NA	NA	20.3	12.02	03/13/2025	10:00

WORKLIST(Hardcopy Internal Chain)

135015

WorkList Name : PH W Q1553 WorkList ID : 188249 Department : Wet-Chemistry Date : 03-13-2025 08:31:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1553-03 -17	FRAC-TANK-FMI120	Water	pH	Cool 4 deg C	PSEG03	I31	03/12/2025	9040C

Date/Time 03/13/25 08:40

Raw Sample Received by: JO YOC

Raw Sample Relinquished by: JO YOC

Date/Time 03/13/25 12:00

Raw Sample Received by: JO YOC

Raw Sample Relinquished by: JO YOC