

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135016

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	03/13/2025	11:40
2	Q1536-01		1	14	100.00	>212.0	>212.0	03/13/2025	12:10
3	Q1536-01DUP		1	14	100.00	>212.0	>212.0	03/13/2025	12:40
4	Q1536-02		1	16	100.00	>212.0	>212.0	03/13/2025	13:10
5	Q1536-03		1	13	100.00	>212.0	>212.0	03/13/2025	13:40
6	Q1553-03		1	16	100.00	>212.0	>212.0	03/13/2025	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB135016

WorkList Name : fp-03-13

WorkList ID : 188234

Department : Wet-Chemistry

Date : 03-13-2025 07:47:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1536-01 B	60425	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/10/2025	1010B
Q1536-02 B	60426	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/10/2025	1010B
Q1536-03 B	60427	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/10/2025	1010B
Q1553-03 D	FRAC-TANK-FMI120	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/12/2025	1010B

Date/Time 03/13/2025 11:30
 Raw Sample Received by: RIM CWS
 Raw Sample Relinquished by: JH WOC

Date/Time 03/13/2025 14:50
 Raw Sample Received by: JH WOC
 Raw Sample Relinquished by: RIM CWS