

LB135019


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4104E



A= 0.0000e+000

B= 2.4640e-004

C= 8.4064e-003

Rho= 0.9997343

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.001	-0.001	-40	0.000	-40					-
0.05	0.050	0.058	0.008	200	0.0 %	200	0				16
0.2	0.200	0.211	0.011	823	0.0 %	823					5
2.5	2.500	2.549	0.049	10310	0.0 %	10310					2
5.0	5.000	4.994	-0.006	20235	0.0 %	20235					0
7.5	7.500	7.317	-0.183	29660	0.0 %	29660					2
10.0	10.000	10.123	0.123	41048	0.0 %	41048					1

LB135019 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Mar 2025 10:53:28

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-40	-	0.0000	245.1	PPB	13 Mar 2025 11:03:16	S	Std
0.05 - 1	50.05	200	-	0.0500	245.1	PPB	13 Mar 2025 11:05:33	S	Std
0.2 - 1	50.2	823	-	0.2000	245.1	PPB	13 Mar 2025 11:07:51	S	Std
2.5 - 1	52.5	10310	-	2.5000	245.1	PPB	13 Mar 2025 11:10:08	S	Std
5.0 - 1	55	20235	-	5.0000	245.1	PPB	13 Mar 2025 11:12:24	S	Std
7.5 - 1	57.5	29660	-	7.5000	245.1	PPB	13 Mar 2025 11:14:42	S	Std
10.0 - 1	510	41048	-	10.0000	245.1	PPB	13 Mar 2025 11:17:00	S	Std
ICV41 - 1	ICV41	16896	4.1716	-245.1		PPB	13 Mar 2025 11:20:13	U	SMPL
ICB41 - 1	ICB41	-219	-0.0456	-245.1		PPB	13 Mar 2025 11:22:28	U	SMPL
CCV06 - 1	CCV06	21575	5.3245	-245.1		PPB	13 Mar 2025 11:24:45	U	SMPL
CCB06 - 1	CCB06	-270	-0.0581	-245.1		PPB	13 Mar 2025 11:27:00	U	SMPL
CRA - 1	CRA	207	0.0594	-245.1		PPB	13 Mar 2025 11:40:23	U	SMPL
HighStd - 1	HighStd	43651	10.7641	-245.1		PPB	13 Mar 2025 11:42:38	U	SMPL
ChkStd - 1	ChkStd	28090	6.9298	-245.1		PPB	13 Mar 2025 11:44:53	U	SMPL
PB167120BL - 1	PBW	-352	-0.0783	-245.1		PPB	13 Mar 2025 11:47:12	U	SMPL
PB167120BS - 1	LCSW	16367	4.0413	-245.1		PPB	13 Mar 2025 11:51:45	U	SMPL
Q1168-10 - 1	LOD-MDL-WATER-04-QT1-2025	120	0.0380	-245.1		PPB	13 Mar 2025 11:56:38	U	SMPL
Q1168-11 - 1	LOQ-WATER-05-QT1-2025	885	0.2265	-245.1		PPB	13 Mar 2025 11:58:54	U	SMPL
Q1491-02 - 1	COMP	588	0.1533	-245.1		PPB	13 Mar 2025 12:01:12	U	SMPL
Q1491-02DUP - 1	COMP DUP	625	0.1624	-245.1		PPB	13 Mar 2025 12:03:28	U	SMPL
Q1491-02MS - 1	COMPMS	14873	3.6731	-245.1		PPB	13 Mar 2025 12:08:35	U	SMPL
CCV07 - 1	CCV07	21266	5.2484	-245.1		PPB	13 Mar 2025 12:10:51	U	SMPL
CCB07 - 1	CCB07	-229	-0.0480	-245.1		PPB	13 Mar 2025 12:13:08	U	SMPL
Q1491-02MSD - 1	COMPMSD	17034	4.2056	-245.1		PPB	13 Mar 2025 12:15:25	U	SMPL
Q1491-02LX5 - 1		-139	-0.0258	-245.1		PPB	13 Mar 2025 12:17:41	U	SMPL
Q1491-02A - 1		16156	3.9893	-245.1		PPB	13 Mar 2025 12:19:59	U	SMPL
CCV08 - 1	CCV08	20342	5.0207	-245.1		PPB	13 Mar 2025 12:22:15	U	SMPL
CCB08 - 1	CCB08	-240	-0.0507	-245.1		PPB	13 Mar 2025 12:24:33	U	SMPL