

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135030

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/14/2025	08:35
2	CAL2	1	Water	NA	NA	20.2	7.00	03/14/2025	08:36
3	CAL3	1	Water	NA	NA	20.3	10.02	03/14/2025	08:40
4	ICV	1	Water	NA	NA	20.3	7.00	03/14/2025	08:45
5	CCV1	1	Water	NA	NA	20.3	2.01	03/14/2025	08:50
6	Q1563-01	1	Water	NA	NA	20.9	5.74	03/14/2025	09:00
7	Q1563-01DUP	1	Water	NA	NA	20.8	5.75	03/14/2025	09:02
8	Q1563-02	1	Water	NA	NA	20.1	7.04	03/14/2025	09:15
9	CCV2	1	Water	NA	NA	20.3	12.02	03/14/2025	09:25

WORKLIST(Hardcopy Internal Chain)

135030

WorkList Name : ph w q1563

WorkList ID : 188268

Department : Wet-Chemistry

Date : 03-14-2025 08:17:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1563-01	17 437	Water	pH	Cool 4 deg C	PSEG03	I51	03/13/2025	9040C
Q1563-02	17 FERNOT-WATER	Water	pH	Cool 4 deg C	PSEG03	I51	03/13/2025	9040C

Date/Time 03/14/25 08:25

Raw Sample Received by: 18 WOC

Raw Sample Relinquished by: CP SM

Date/Time 03/14/25

Raw Sample Received by: CP SM

Raw Sample Relinquished by: 18 WOC