

L3135034


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

5453E



A= 0.0000e+000

B= 1.8355e-004

C= 4.3543e-002

Rho= 0.9997643

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	0.100	0.100	305	0.000	305	0				1
0.2	0.200	0.239	0.039	1065	0.0 %	1065					20
2.5	2.500	2.352	-0.148	12575	0.0 %	12575					6
5.0	5.000	4.953	-0.047	26750	0.0 %	26750					1
7.5	7.500	7.503	0.003	40639	0.0 %	40639					0
10.0	10.000	10.054	0.054	54538	0.0 %	54538					1

LB135034 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Mar 2025 10:14:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	305	-	0.0000	7470A	PPB	14 Mar 2025 10:33:32	S	Std
0.2 - 1	50-2	1065	-	0.2000	7470A	PPB	14 Mar 2025 10:35:49	S	Std
2.5 - 1	52.5	12575	-	2.5000	7470A	PPB	14 Mar 2025 10:38:06	S	Std
5.0 - 1	55	26750	-	5.0000	7470A	PPB	14 Mar 2025 10:43:00	S	Std
7.5 - 1	57.5	40639	-	7.5000	7470A	PPB	14 Mar 2025 10:48:27	S	Std
10.0 - 1	511	54538	-	10.0000	7470A	PPB	14 Mar 2025 10:55:43	S	Std
ICV43 - 1	ICV43	19830	3.6833	-	7470A	PPB	14 Mar 2025 11:11:28	U	SMPL
ICB43 - 1	ICB43	274	0.0938	-	7470A	PPB	14 Mar 2025 11:13:45	U	SMPL
CCV12 - 1	CCV12	26114	4.8367	-	7470A	PPB	14 Mar 2025 11:16:02	U	SMPL
CCB12 - 1	CCB12	-60	0.0325	-	7470A	PPB	14 Mar 2025 11:18:17	U	SMPL
CRA - 1	CRA	1004	0.2278	-	7470A	PPB	14 Mar 2025 11:20:34	U	SMPL
HighStd - 1	HighStd	54498	10.0464	-	7470A	PPB	14 Mar 2025 11:22:49	U	SMPL
ChkStd - 1	ChkStd	37207	6.8727	-	7470A	PPB	14 Mar 2025 11:25:04	U	SMPL
PB167121BL - 1	PBW	-361	-0.0227	-	7470A	PPB	14 Mar 2025 11:27:23	U	SMPL
Q1547-05 - 1	OR-620-JB-COMP-01	-171	0.0122	-	7470A	PPB	14 Mar 2025 11:34:56	U	SMPL
Q1547-05DUP - 1	OR-620-JB-COMP-01DUP	93	0.0606	-	7470A	PPB	14 Mar 2025 11:37:14	U	SMPL
Q1547-05MS - 1	OR-620-JB-COMP-01MS	15220	2.8371	-	7470A	PPB	14 Mar 2025 11:39:31	U	SMPL
Q1547-05MSD - 1	OR-620-JB-COMP-01MSD	14559	2.7158	-	7470A	PPB	14 Mar 2025 11:41:48	U	SMPL
Q1547-10 - 1	OR-620-JB-COMP-02	-18	0.0402	-	7470A	PPB	14 Mar 2025 11:44:04	U	SMPL
CCV13 - 1	CCV13	26117	4.8372	-	7470A	PPB	14 Mar 2025 11:46:21	U	SMPL
CCB13 - 1	CCB13	219	0.0837	-	7470A	PPB	14 Mar 2025 11:48:37	U	SMPL
Q1552-05 - 1	NB-418-JB-COMP-01	164	0.0736	-	7470A	PPB	14 Mar 2025 11:50:54	U	SMPL
PB167103TB - 1	PB167103TB	94	0.0608	-	7470A	PPB	14 Mar 2025 11:53:09	U	SMPL
Q1547-05LX5 - 1		11	0.0456	-	7470A	PPB	14 Mar 2025 11:55:25	U	SMPL
Q1547-05A - 1		24239	4.4925	-	7470A	PPB	14 Mar 2025 11:57:41	U	SMPL
PB167121BS - 1	LCSW	24164	4.4787	-	7470A	PPB	14 Mar 2025 11:59:57	U	SMPL
CCV14 - 1	CCV14	24389	4.5200	-	7470A	PPB	14 Mar 2025 12:02:15	U	SMPL
CCB14 - 1	CCB14	87	0.0595	-	7470A	PPB	14 Mar 2025 12:04:34	U	SMPL