

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135039

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/14/2025	14:40
2	CAL2	1	Water	NA	NA	20.2	7.00	03/14/2025	14:41
3	CAL3	1	Water	NA	NA	20.2	10.02	03/14/2025	14:45
4	ICV	1	Water	NA	NA	20.1	6.99	03/14/2025	14:50
5	CCV1	1	Water	NA	NA	20.2	2.01	03/14/2025	15:00
6	Q1569-01	1	Solid	20.02	20	21.4	6.67	03/14/2025	15:10
7	Q1569-04	1	Solid	20.03	20	21.5	10.20	03/14/2025	15:20
8	Q1574-01	1	Solid	20.03	20	22.2	6.76	03/14/2025	15:35
9	Q1574-01DUP	1	Solid	20.04	20	22.3	6.77	03/14/2025	15:36
10	CCV2	1	Water	NA	NA	20.2	12.02	03/14/2025	15:40

WORKLIST(Hardcopy Internal Chain)

135039

WorkList Name : ph q1574

WorkList ID : 188292

Department : Wet-Chemistry

Date : 03-14-2025 14:23:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1569-01	D TAP-IDW-SOIL-031325-01	Solid	pH	Cool 4 deg C	WEST04	I33	03/13/2025	9045D
Q1569-04	D TAP-IDW-SOIL-031325-02	Solid	pH	Cool 4 deg C	WEST04		03/13/2025	9045D
Q1574-01	WC1	Solid	pH	Cool 4 deg C	GENV01	I41	03/12/2025	9045D

Date/Time 03/14/25 14:30

Raw Sample Received by: JD WOC

Raw Sample Relinquished by: CP SM

Date/Time 03/14/25

Raw Sample Received by: CP SM

Raw Sample Relinquished by: JD WOC

14:00