

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135050

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/17/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	03/17/2025	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	03/17/2025	08:19
4	ICV	1	Water	NA	NA	20.2	7.00	03/17/2025	08:22
5	CCV1	1	Water	NA	NA	20.3	2.01	03/17/2025	08:25
6	Q1577-03	1	Water	NA	NA	20.3	8.03	03/17/2025	08:35
7	Q1577-03DUP	1	Water	NA	NA	20.4	8.04	03/17/2025	08:36
8	CCV2	1	Water	NA	NA	20.3	12.02	03/17/2025	08:40

WORKLIST(Hardcopy Internal Chain)

VB 135050

WorkList Name : ph w q1577 WorkList ID : 188296 Department : Wet-Chemistry Date : 03-17-2025 07:58:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1577-03	MOO-25-0072	Water	pH	Cool 4 deg C	PSEG03	I41	03/14/2025	9040C

Date/Time 03/17/25 08:05
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 03/17/25 10:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]