

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135052

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	03/17/2025	08:30
2	Q1563-01		1	15	100.00	>212.0	>212.0	03/17/2025	09:00
3	Q1563-01DUP		1	15	100.00	>212.0	>212.0	03/17/2025	09:30
4	Q1563-02		1	12	100.00	>212.0	>212.0	03/17/2025	10:00
5	Q1577-03		1	15	100.00	>212.0	>212.0	03/17/2025	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16135052

WorkList Name : FP-3-14

WorkList ID : 188291

Department : Wet-Chemistry

Date : 03-14-2025 14:22:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1563-01D	437	Water	Flash Point	Cool 4 deg C	PSEG03	I51	03/13/2025	1010B
Q1563-02D	FERNOT-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	I51	03/13/2025	1010B
Q1577-03D	MOO-25-0072	Water	Flash Point	Cool 4 deg C	PSEG03	I41	03/14/2025	1010B

Date/Time 03/17/2025 0820
 Raw Sample Received by: RM (w/17/2025)
 Raw Sample Relinquished by: JH (w/2025)

Date/Time 03/17/2025 11.10
 Raw Sample Received by: JH (w/2025)
 Raw Sample Relinquished by: RM (w/2025)