

Analytical Summary Report

Analysis Method: 120.1
 Parameter: Conductance
 Run Number: LB135058

ANALYST: Niha
 SUPERVISOR REVIEW BY: Iwona
 Meter ID: WC CONDUCTANCE METER-1
 Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP111962

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB135058BL		1	22.5	0.83	0.87	03/13/2025	13:00
2	LB135058BS	1413	1	22.7	1347.00	1408.89	03/13/2025	13:05
3	Q1505-08		1	22.8	322.00	336.12	03/13/2025	13:10
4	Q1505-08DUP		1	22.7	320.00	334.70	03/13/2025	13:15

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

WORKLIST(Hardcopy Internal Chain)

LB135058

WorkList Name : CONDUCTANCE PT -03

WorkList ID : 188259

Department : Wet-Chemistry

Date : 03-13-2025 11:37:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1505-08	PT-COND-WP	Water	Conductance	Cool 4 deg C	ALLI03	QA Of	03/03/2025	120.1

Date/Time 03.13.2025 11:45
 Raw Sample Received by: NF(WC)
 Raw Sample Relinquished by: SJ(QA0)

Date/Time NA
 Raw Sample Received by: "
 Raw Sample Relinquished by: "