

6135060

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

3/17/2025 15:56

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.983	0.0	0.100	
ICB1	-0.034	0.0	0.004	
CCV1	0.992	0.0	0.101	
CCB1	-0.031	0.0	0.005	
PB167174BL	-0.032	0.0	0.004	
PB167174BS	0.991	0.0	0.101	
Q1505-26	1.272	0.0	0.127	
CCV2	0.999	0.0	0.101	
CCB2	-0.032	0.0	0.004	

N 9
Mean 0.567
SD 0.5760
CV% 101.50

Aquakem v. 7.2AQ1

Results from time period:

Mon Mar 17 14:07:03 2025

Mon Mar 17 15:33:28 2025

Sample Id	Sam/Ctr/c	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0164	mg/l	3/17/2025 14:07:03	
0.05PPM	A	Phenolics-	P	0.0529	mg/l	3/17/2025 14:07:04	
0.1PPM	A	Phenolics-	P	0.1017	mg/l	3/17/2025 14:07:05	
0.25PPM	A	Phenolics-	P	0.2593	mg/l	3/17/2025 14:07:06	
0.50PPM	A	Phenolics-	P	0.5081	mg/l	3/17/2025 14:07:07	
1.0PPM	A	Phenolics-	P	0.9956	mg/l	3/17/2025 14:07:08	
2.0PPM	A	Phenolics-	P	1.9989	mg/l	3/17/2025 14:07:09	
ICV1	S	Phenolics-	P	0.9832	mg/l	3/17/2025 15:23:15	
ICB1	S	Phenolics-	P	-0.0338	mg/l	3/17/2025 15:23:18	
CCV1	S	Phenolics-	P	0.9917	mg/l	3/17/2025 15:23:20	
CCB1	S	Phenolics-	P	-0.0312	mg/l	3/17/2025 15:23:21	
PB167174BL	S	Phenolics-	P	-0.0323	mg/l	3/17/2025 15:23:23	
PB167174BS	S	Phenolics-	P	0.9906	mg/l	3/17/2025 15:23:25	
Q1505-26	S	Phenolics-	P	1.272	mg/l	3/17/2025 15:33:20	
CCV2	S	Phenolics-	P	0.9991	mg/l	3/17/2025 15:33:26	
CCB2	S	Phenolics-	P	-0.0322	mg/l	3/17/2025 15:33:27	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM Instrument ID : Konelab

3/17/2025 14:09


Test Phenolics-

Accepted 3/17/2025 14:09

Factor 10.64
Bias 0.007

Coeff. of det. 0.999856

Errors



	Calibrator	Response	Calc. con.	Conc.	^Q Errors
1	0.00PPM	0.006	-0.0164	0.0000	-
2	Phenol-2	0.012	0.0529	0.0500	5.8
3	Phenol-2	0.017	0.1017	0.1000	1.7
4	Phenol-2	0.032	0.2593	0.2500	3.7
5	Phenol-2	0.055	0.5081	0.5000	1.6
6	Phenol-2	0.101	0.9956	1.0000	-0.4
7	Phenol-2	0.195	1.9989	2.0000	-0.1

03/17/2025
RM