

LB135064

7470A

INSTRUMENT ID: LV1

Linear

 μ Abs.:

40917

A= 0.0000e+000

B= 2.4024e-004

C= -3.2981e-002

Rho= 0.9985198

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.020	-0.020	55	0.000	55	0				0.00
0.2	0.200	0.172	-0.028	855	0.0 %	855					-14
2.5	2.500	2.636	0.136	11108	0.0 %	11108					5
5.0	5.000	4.762	-0.238	19960	0.0 %	19960					-5
7.5	7.500	7.853	0.353	32824	0.0 %	32824					5
10.0	10.000	9.797	-0.203	40917	0.0 %	40917					2

LB135064 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Mar 2025 07:56:17

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	SD	55	-	0.0000	7470A	PPB	18 Mar 2025 09:34:39	S	Std
0.2 - 1	SD-2	855	-	0.2000	7470A	PPB	18 Mar 2025 09:36:56	S	Std
2.5 - 1	SD-5	11108	-	2.5000	7470A	PPB	18 Mar 2025 09:39:14	S	Std
5.0 - 1	SD	19960	-	5.0000	7470A	PPB	18 Mar 2025 09:41:31	S	Std
7.5 - 1	SD-5	32824	-	7.5000	7470A	PPB	18 Mar 2025 09:49:56	S	Std
10.0 - 1	SD	40917	-	10.0000	7470A	PPB	18 Mar 2025 09:52:11	S	Std
ICV46 - 1	ICV46	16548	3.9425	-	7470A	PPB	18 Mar 2025 09:55:05	U	SMPL
ICB46 - 1	ICB46	-227	-0.0875	-	7470A	PPB	18 Mar 2025 09:57:22	U	SMPL
CCV20 - 1	CCV20	22482	5.3681	-	7470A	PPB	18 Mar 2025 09:59:38	U	SMPL
CCB20 - 1	CCB20	-43	-0.0433	-	7470A	PPB	18 Mar 2025 10:04:36	U	SMPL
CRA - 1	CRA	808	0.1611	-	7470A	PPB	18 Mar 2025 10:06:51	U	SMPL
HighStd - 1	HighStd	42000	10.0571	-	7470A	PPB	18 Mar 2025 10:09:06	U	SMPL
ChkStd - 1	ChkStd	29904	7.1511	-	7470A	PPB	18 Mar 2025 10:11:22	U	SMPL
PB167166BL - 1	PBW	-3	-0.0337	-	7470A	PPB	18 Mar 2025 10:16:17	U	SMPL
PB167166BS - 1	LCSW	19273	4.5972	-	7470A	PPB	18 Mar 2025 10:18:34	U	SMPL
Q1568-04 - 1	JC-03-03132025	-244	-0.0916	-	7470A	PPB	18 Mar 2025 10:20:52	U	SMPL
Q1568-04DUP - 1	JC-03-03132025DUP	-3	-0.0337	-	7470A	PPB	18 Mar 2025 10:23:10	U	SMPL
Q1568-04MS - 1	JC-03-03132025MS	16522	3.9363	-	7470A	PPB	18 Mar 2025 10:25:27	U	SMPL
Q1568-04MSD - 1	JC-03-03132025MSD	16341	3.8928	-	7470A	PPB	18 Mar 2025 10:27:44	U	SMPL
Q1569-02 - 1	TAP-IDW-SOIL-031325-01	-87	-0.0539	-	7470A	PPB	18 Mar 2025 10:30:02	U	SMPL
CCV21 - 1	CCV21	20408	4.8698	-	7470A	PPB	18 Mar 2025 10:35:02	U	SMPL
CCB21 - 1	CCB21	47	-0.0217	-	7470A	PPB	18 Mar 2025 10:41:17	U	SMPL
Q1569-05 - 1	TAP-IDW-SOIL-031325-02	42	-0.0229	-	7470A	PPB	18 Mar 2025 10:43:32	U	SMPL
Q1574-02 - 1	WC1	2309	0.5217	-	7470A	PPB	18 Mar 2025 10:45:47	U	SMPL
PB167133TB - 1	PB167133TB	-66	-0.0488	-	7470A	PPB	18 Mar 2025 10:48:03	U	SMPL
Q1568-04LX5 - 1		25	-0.0270	-	7470A	PPB	18 Mar 2025 10:50:19	U	SMPL
Q1568-04A - 1		16440	3.9166	-	7470A	PPB	18 Mar 2025 10:52:35	U	SMPL
CCV22 - 1	CCV22	21462	5.1230	-	7470A	PPB	18 Mar 2025 10:54:51	U	SMPL
CCB22 - 1	CCB22	4	-0.0320	-	7470A	PPB	18 Mar 2025 10:59:43	U	SMPL