

TitriSoft Report

Ident	Amount	DateTime	Alkalinity	Final pH (sample)	Initial pH (sample)	slope	offset
LB135065BLW	100	18/03/2025 09:30	0.00	3.87	3.91	102.5	6.86
LB135065BSW	100	18/03/2025 09:35	44.96	4.40	8.70	102.5	6.86
Q1505-05	50	18/03/2025 09:40	344.44	4.46	8.32	102.5	6.86
Q1525-01	50	18/03/2025 09:45	153.62	4.46	7.68	102.5	6.86
Q1525-01DUP	50	18/03/2025 09:50	153.68	4.45	7.68	102.5	6.86

03/18/2025

03.18.2025

NF

WORKLIST(Hardcopy Internal Chain)

LB135065

WorkList Name : alkalinity-3-10

WorkList ID : 188153

Department : Wet-Chemistry

Date : 03-10-2025 09:18:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1505-05	PT-MIN1-WP	Water	Alkalinity	Cool 4 deg C	ALLI03	QA Of	03/03/2025	SM2320 B
Q1525-01	MW10	Water	Alkalinity	Cool 4 deg C	GENV01	I31	03/06/2025	SM2320 B

Date/Time 03.18.2025 , 08:50
 Raw Sample Received by: NFWC
 Raw Sample Relinquished by: NFWC

Date/Time 03.18.2025 , 10:00
 Raw Sample Received by: NFWC
 Raw Sample Relinquished by: NFWC