

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB135087

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP112352
Chlorine Calibration std, 0.1ppm	WP112347
Chlorine Calibration std, 0.2ppm	WP112348
Chlorine Calibration std, 0.8ppm	WP112349
Chlorine Calibration std, 0.0ppm	WP112346
Chlorine Calibration std, 1.6ppm	WP112350
Residual Chlorine Calibration and CCV std, 0	WP112351
Total Chlorine Powder Pillows	W3147

Intercept: 0.0014

Slope: 0.9327

Regression: 0.999572

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	0.00		03/19/2025	10:45
2	CAL2	0.1	1	0.000	0.090	0.090	0.10	-5	03/19/2025	10:48
3	CAL3	0.2	1	0.000	0.180	0.180	0.19	-4.5	03/19/2025	10:51
4	CAL4	0.4	1	0.000	0.370	0.370	0.40	-1.3	03/19/2025	10:54
5	CAL5	0.8	1	0.000	0.780	0.780	0.84	4.4	03/19/2025	10:57
6	CAL6	1.6	1	0.000	1.480	1.480	1.59	-0.9	03/19/2025	11:00

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB135087

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3800	0.3800	0.4100	03/19/2025	11:03
2	ICB				1	0.0000	0.0000	0.0000	-0.0080	03/19/2025	11:06
3	CCV1			0.4	1	0.0000	0.3800	0.3800	0.4100	03/19/2025	11:09
4	CCB1				1	0.0000	0.0000	0.0000	-0.0080	03/19/2025	11:12
5	LB135087BL	50	50		1	0.0000	0.0000	0.0000	-0.0080	03/19/2025	11:15
6	LB135087BS	50	50	0.4	1	0.0000	0.3900	0.3900	0.4210	03/19/2025	11:18
7	Q1505-19	50	50		1	0.0000	1.0800	1.0800	1.1560	03/19/2025	11:21
8	CCV2			0.4	1	0.0000	0.3600	0.3600	0.3880	03/19/2025	11:24
9	CCB2				1	0.0000	0.0000	0.0000	-0.0080	03/19/2025	11:27

WORKLIST(Hardcopy Internal Chain)

LB135087

WorkList Name :	RESIDUAL CHLORINE F	WorkList ID :	188369	Department :	Wet-Chemistry	Date :	03-19-2025 08:50:40	
Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1505-19	PT-CL-WP	Water	Residual Chlorine	Cool 4 deg C	ALLI03	QA Of	03/03/2025	SM4500 Cl G

Date/Time 03.19.2025 10:00
Raw Sample Received by: NF(wc)
Raw Sample Relinquished by: 12(wc)

Date/Time NA
Raw Sample Received by: "
Raw Sample Relinquished by: y