

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135096

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/19/2025	16:30
2	CAL2	1	Water	NA	NA	20.3	7.00	03/19/2025	16:31
3	CAL3	1	Water	NA	NA	20.3	10.02	03/19/2025	16:33
4	ICV	1	Water	NA	NA	20.3	7.02	03/19/2025	16:35
5	CCV1	1	Water	NA	NA	20.2	2.01	03/19/2025	16:40
6	Q1604-02	1	Water	NA	NA	20.5	7.11	03/19/2025	16:45
7	Q1604-02DUP	1	Water	NA	NA	20.6	7.12	03/19/2025	16:46
8	Q1606-11	1	Water	NA	NA	20.1	6.85	03/19/2025	16:50
9	CCV2	1	Water	NA	NA	20.3	12.02	03/20/2025	17:00

WORKLIST(Hardcopy Internal Chain)

VB135096

WorkList Name : ph q1604

WorkList ID : 188377

Department : Wet-Chemistry

Date : 03-19-2025 14:03:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1604-02	R FRAC-TANK-N45878	Water	pH	Cool 4 deg C	PSEG03	I41	03/19/2025	9040C
Q1606-11	R N48965	Water	pH	Cool 4 deg C	PSEG03	I41	03/19/2025	9040C

Date/Time 03/19/25 16:15

Raw Sample Received by: SB WCC

Raw Sample Relinquished by: OP SN

Date/Time 03/19/25 18:00

Raw Sample Received by: OP SN

Raw Sample Relinquished by: SB WCC