

LB135100


7470A

INSTRUMENT ID: CVD

Linear

 μ Abs.:

45893



A= 0.0000e+000

B= 2.2234e-004

C= -1.2078e-001

Rho= 0.9993955

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.051	-0.051	316	0.000	316	0				1
0.2	0.200	0.205	0.005	1467	0.0 %	1467					2
2.5	2.500	2.722	0.222	12786	0.0 %	12786					9
5.0	5.000	4.814	-0.186	22196	0.0 %	22196					1
7.5	7.500	7.426	-0.074	33943	0.0 %	33943					1
10.0	10.000	10.083	0.083	45893	0.0 %	45893					1

LB135100 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Mar 2025 07:50:36

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	316	-	0.0000	7470A	PPB	20 Mar 2025 08:10:39	S	Std
0.2 - 1	50.2	1467	-	0.2000	7470A	PPB	20 Mar 2025 08:12:56	S	Std
2.5 - 1	52.15	12786	-	2.5000	7470A	PPB	20 Mar 2025 08:15:13	S	Std
5.0 - 1	55	22196	-	5.0000	7470A	PPB	20 Mar 2025 08:17:30	S	Std
7.5 - 1	57.5	33943	-	7.5000	7470A	PPB	20 Mar 2025 08:19:48	S	Std
10.0 - 1	514	45893	-	10.0000	7470A	PPB	20 Mar 2025 08:26:24	S	Std
ICV49 - 1	ICV49	19050	4.1147	-	7470A	PPB	20 Mar 2025 08:32:23	U	SMPL
ICB49 - 1	ICB49	221	-0.0716	-	7470A	PPB	20 Mar 2025 08:34:38	U	SMPL
CCV28 - 1	CCV28	24197	5.2591	-	7470A	PPB	20 Mar 2025 08:39:28	U	SMPL
CCB28 - 1	CCB28	327	-0.0481	-	7470A	PPB	20 Mar 2025 08:44:22	U	SMPL
CRA - 1	CRA	1433	0.1978	-	7470A	PPB	20 Mar 2025 08:52:14	U	SMPL
HighStd - 1	HighStd	47186	10.3704	-	7470A	PPB	20 Mar 2025 08:58:37	U	SMPL
ChkStd - 1	ChkStd	34066	7.4533	-	7470A	PPB	20 Mar 2025 09:03:29	U	SMPL
PB167213BL - 1	PBW	418	-0.0278	-	7470A	PPB	20 Mar 2025 09:08:22	U	SMPL
PB167213BS - 1	LCSW	15578	3.3428	-	7470A	PPB	20 Mar 2025 09:10:37	U	SMPL
Q1590-01 - 1	3794	-111	-0.1455	-	7470A	PPB	20 Mar 2025 09:12:54	U	SMPL
Q1592-02 - 1	OILY-DEBRIS-COMP	56	-0.1083	-	7470A	PPB	20 Mar 2025 09:15:11	U	SMPL
Q1597-02 - 1	1-CONCRETE-SLAB	581	0.0084	-	7470A	PPB	20 Mar 2025 09:17:28	U	SMPL
Q1597-04 - 1	2-CONCRETE-SLAB	52	-0.1092	-	7470A	PPB	20 Mar 2025 09:19:45	U	SMPL
Q1597-06 - 1	3-CONCRETE-SLAB	38	-0.1123	-	7470A	PPB	20 Mar 2025 09:22:01	U	SMPL
CCV29 - 1	CCV29	21430	4.6439	-	7470A	PPB	20 Mar 2025 09:26:51	U	SMPL
CCB29 - 1	CCB29	405	-0.0307	-	7470A	PPB	20 Mar 2025 09:31:55	U	SMPL
Q1597-06DUP - 1	3-CONCRETE-SLABDUP	63	-0.1068	-	7470A	PPB	20 Mar 2025 09:34:09	U	SMPL
Q1597-06MS - 1	3-CONCRETE-SLABMS	17291	3.7236	-	7470A	PPB	20 Mar 2025 09:36:25	U	SMPL
Q1597-06MSD - 1	3-CONCRETE-SLABMSD	18814	4.0623	-	7470A	PPB	20 Mar 2025 09:41:49	U	SMPL
PB167220BL - 1	PBW	271	-0.0605	-	7470A	PPB	20 Mar 2025 09:44:04	U	SMPL
PB167220BS - 1	LCSW	20897	4.5254	-	7470A	PPB	20 Mar 2025 09:46:22	U	SMPL
Q1592-03 - 1	257952	57	-0.1081	-	7470A	PPB	20 Mar 2025 09:48:39	U	SMPL
Q1592-03DUP - 1	257952DUP	189	-0.0788	-	7470A	PPB	20 Mar 2025 09:50:57	U	SMPL
Q1592-03MS - 1	257952MS	15410	3.3054	-	7470A	PPB	20 Mar 2025 09:53:14	U	SMPL
Q1592-03MSD - 1	257952MSD	19111	4.1283	-	7470A	PPB	20 Mar 2025 10:00:59	U	SMPL
Q1593-01 - 1	258829	6648	1.3573	-	7470A	PPB	20 Mar 2025 10:03:14	U	SMPL
CCV30 - 1	CCV30	24370	5.2976	-	7470A	PPB	20 Mar 2025 10:05:32	U	SMPL
CCB30 - 1	CCB30	439	-0.0232	-	7470A	PPB	20 Mar 2025 10:07:49	U	SMPL
Q1594-01 - 1	GAS-CAM-40615	1115	0.1271	-	7470A	PPB	20 Mar 2025 10:10:07	U	SMPL
Q1604-02 - 1	FRAC-TANK-N45878	1632	0.2421	-	7470A	PPB	20 Mar 2025 10:12:23	U	SMPL
PB167193TB - 1	PB167193TB	539	-0.0009	-	7470A	PPB	20 Mar 2025 10:19:59	U	SMPL
Q1597-06LX5 - 1		4	-0.1199	-	7470A	PPB	20 Mar 2025 10:22:14	U	SMPL
Q1597-06A - 1		18912	4.0841	-	7470A	PPB	20 Mar 2025 10:24:30	U	SMPL
Q1592-03LX5 - 1		-154	-0.1550	-	7470A	PPB	20 Mar 2025 10:26:46	U	SMPL
Q1592-03A - 1		18979	4.0989	-	7470A	PPB	20 Mar 2025 10:29:06	U	SMPL
CCV31 - 1	CCV31	22697	4.9256	-	7470A	PPB	20 Mar 2025 10:31:26	U	SMPL
CCB31 - 1	CCB31	462	-0.0181	-	7470A	PPB	20 Mar 2025 10:33:45	U	SMPL