

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135112

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/20/2025	15:10
2	CAL2	1	Water	NA	NA	20.3	7.00	03/20/2025	15:11
3	CAL3	1	Water	NA	NA	20.3	10.02	03/20/2025	15:15
4	ICV	1	Water	NA	NA	20.3	7.02	03/20/2025	15:20
5	CCV1	1	Water	NA	NA	20.3	2.01	03/20/2025	15:25
6	Q1618-01	1	Water	NA	NA	20.5	6.86	03/20/2025	15:35
7	Q1618-01DUP	1	Water	NA	NA	20.6	6.87	03/20/2025	15:36
8	Q1618-02	1	Water	NA	NA	20.6	10.88	03/20/2025	15:50
9	CCV2	1	Water	NA	NA	20.2	12.02	03/20/2025	15:55

WORKLIST(Hardcopy Internal Chain)

UR 135112

WorkList Name : ph w q1618

WorkList ID : 188417

Department : Wet-Chemistry

Date : 03-20-2025 14:49:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1618-01 E	TP20250320-01	Water	pH	Cool 4 deg C	PSEG03	I31	03/20/2025	9040C
Q1618-02 E	TP20250320-02	Water	pH	Cool 4 deg C	PSEG03	I31	03/20/2025	9040C

Date/Time 03/20/25 15:00

Raw Sample Received by: af wcy

Raw Sample Relinquished by: af sm

Date/Time 03/20/25 18:00

Raw Sample Received by: af sm

Raw Sample Relinquished by: af wcy