

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135122

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	03/21/2025	09:40
2	Q1591-04		1	13	100.00	>212.0	>212.0	03/21/2025	10:10
3	Q1591-04DUP		1	13	100.00	>212.0	>212.0	03/21/2025	10:40
4	Q1591-05		1	15	100.00	>212.0	>212.0	03/21/2025	11:10
5	Q1591-06		1	15	100.00	>212.0	>212.0	03/21/2025	11:40
6	Q1604-02		1	16	100.00	>212.0	>212.0	03/21/2025	12:10
7	Q1606-11		1	14	100.00	>212.0	>212.0	03/21/2025	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66135122

WorkList Name : FP-3-19

WorkList ID : 188390

Department : Wet-Chemistry

Date : 03-19-2025 16:04:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1591-04 <i>B</i>	GAS-TRE-25-0026	Water	Flash Point	Cool 4 deg C	PSEG03	I51	03/17/2025	1010B
Q1591-05 <i>B</i>	GAS-TRE-25-0027	Water	Flash Point	Cool 4 deg C	PSEG03	I51	03/17/2025	1010B
Q1591-06 <i>B</i>	GAS-TRE-25-0028	Water	Flash Point	Cool 4 deg C	PSEG03	I51	03/17/2025	1010B
Q1604-02 <i>U1</i>	FRAC-TANK-N45878	Water	Flash Point	Cool 4 deg C	PSEG03	I41	03/19/2025	1010B
Q1606-11 <i>O</i>	N48965	Water	Flash Point	Cool 4 deg C	PSEG03	I41	03/19/2025	1010B

Date/Time 03/21/2025 09:20
 Raw Sample Received by: 12(21)
 Raw Sample Relinquished by: 28(21)

Date/Time 03/21/2025 13:25
 Raw Sample Received by: 28(21)
 Raw Sample Relinquished by: 12(21)