

LB135123


7471B

INSTRUMENT ID: CVJ

Linear

μ Abs.:

57686



A= 0.0000e+000

B= 1.7231e-004

C= 3.9604e-002

Rho= 0.9997964

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.038	0.038	-12	0.000	-12					-
0.2	0.200	0.220	0.020	1049	0.0 %	1049					10
2.5	2.500	2.475	-0.025	14135	0.0 %	14135					-1
5.0	5.000	4.870	-0.130	28035	0.0 %	28035					-3
7.5	7.500	7.617	0.117	43973	0.0 %	43973					2
10.0	10.000	9.980	-0.020	57686	0.0 %	57686					0

LB135123 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Mar 2025 10:23:38

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-12	-	0.0000	7471B	PPB	21 Mar 2025 10:26:37	S	Std
0.2 - 1	50.2	1049	-	0.2000	7471B	PPB	21 Mar 2025 10:28:54	S	Std
2.5 - 1	52.5	14135	-	2.5000	7471B	PPB	21 Mar 2025 10:31:11	S	Std
5.0 - 1	55	28035	-	5.0000	7471B	PPB	21 Mar 2025 10:33:28	S	Std
7.5 - 1	57.5	43973	-	7.5000	7471B	PPB	21 Mar 2025 10:35:47	S	Std
10.0 - 1	516	57686	-	10.0000	7471B	PPB	21 Mar 2025 10:38:06	S	Std
ICV52 - 1	ICV52	24987	4.3452	-	7471B	PPB	21 Mar 2025 10:41:01	U	SMPL
ICB52 - 1	ICB52	-362	-0.0228	-	7471B	PPB	21 Mar 2025 10:43:17	U	SMPL
CCV37 - 1	CCV37	28530	4.9557	-	7471B	PPB	21 Mar 2025 10:45:36	U	SMPL
CCB37 - 1	CCB37	-421	-0.0329	-	7471B	PPB	21 Mar 2025 10:47:52	U	SMPL
CRA - 1	CRA	1000	0.2119	-	7471B	PPB	21 Mar 2025 10:50:10	U	SMPL
HighStd - 1	HighStd	56984	9.8587	-	7471B	PPB	21 Mar 2025 10:52:25	U	SMPL
ChkStd - 1	ChkStd	38076	6.6006	-	7471B	PPB	21 Mar 2025 10:54:42	U	SMPL
PB167246BL - 1	PB167246BL	-455	-0.0388	-	7471B	PPB	21 Mar 2025 10:57:07	U	SMPL
PB167246BS - 1	PB167246BS	24038	4.1817	-	7471B	PPB	21 Mar 2025 10:59:31	U	SMPL
Q1606-01 - 1	CHRT24743	47884	8.2907	-	7471B	PPB	21 Mar 2025 11:01:47	U	SMPL
Q1606-03 - 1	RBR251346	39551	6.8548	-	7471B	PPB	21 Mar 2025 11:04:07	U	SMPL
Q1606-05 - 1	RT4534	16152	2.8228	-	7471B	PPB	21 Mar 2025 11:06:37	U	SMPL
Q1606-07 - 1	RT3025	24166	4.2037	-	7471B	PPB	21 Mar 2025 11:09:04	U	SMPL
Q1606-09 - 1	CHRT28607	60798	10.5159	-	7471B	PPB	21 Mar 2025 11:11:22	U	SMPL
CCV38 - 1	CCV38	26687	4.6381	-	7471B	PPB	21 Mar 2025 11:13:42	U	SMPL
CCB38 - 1	CCB38	-450	-0.0379	-	7471B	PPB	21 Mar 2025 11:16:14	U	SMPL
Q1607-01 - 1	OR-03-031925	7070	1.2579	-	7471B	PPB	21 Mar 2025 11:18:32	U	SMPL
Q1610-02 - 1	SOIL-PILE	2281	0.4327	-	7471B	PPB	21 Mar 2025 11:20:47	U	SMPL
Q1611-01 - 1	RBR200030	1068	0.2236	-	7471B	PPB	21 Mar 2025 11:23:05	U	SMPL
Q1611-01DUP - 1	RBR200030DUP	1051	0.2207	-	7471B	PPB	21 Mar 2025 11:25:21	U	SMPL
Q1611-01MS - 1	RBR200030MS	27056	4.7017	-	7471B	PPB	21 Mar 2025 11:27:37	U	SMPL
Q1611-01MSD - 1	RBR200030MSD	23640	4.1131	-	7471B	PPB	21 Mar 2025 11:29:54	U	SMPL
Q1614-01 - 1	TR-04-032025	97	0.0563	-	7471B	PPB	21 Mar 2025 11:32:14	U	SMPL
Q1619-05 - 1	TP-1	8962	1.5839	-	7471B	PPB	21 Mar 2025 11:34:34	U	SMPL
Q1619-07 - 1	TP-2	1319	0.2669	-	7471B	PPB	21 Mar 2025 11:36:51	U	SMPL
Q1619-09 - 1	TP-3	7642	1.3564	-	7471B	PPB	21 Mar 2025 11:39:09	U	SMPL
CCV39 - 1	CCV39	26852	4.6666	-	7471B	PPB	21 Mar 2025 11:41:25	U	SMPL
CCB39 - 1	CCB39	-405	-0.0302	-	7471B	PPB	21 Mar 2025 11:43:43	U	SMPL
Q1619-11 - 1	TP-4	10207	1.7984	-	7471B	PPB	21 Mar 2025 11:46:01	U	SMPL
Q1619-13 - 1	TP-5	2951	0.5481	-	7471B	PPB	21 Mar 2025 11:48:17	U	SMPL
Q1611-01LX5 - 1		108	0.0582	-	7471B	PPB	21 Mar 2025 11:50:34	U	SMPL
Q1611-01A - 1		24729	4.3007	-	7471B	PPB	21 Mar 2025 11:52:51	U	SMPL
Q1606-09DLX2 - 1		31039	5.3880	-	7471B	PPB	21 Mar 2025 11:55:07	U	SMPL
CCV40 - 1	CCV40	26722	4.6442	-	7471B	PPB	21 Mar 2025 11:57:26	U	SMPL
CCB40 - 1	CCB40	-391	-0.0278	-	7471B	PPB	21 Mar 2025 11:59:48	U	SMPL