

LB135130


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40401



A= 0.0000e+000

B= 2.4409e-004 slope

C= -5.5250e-002 y-intercept

Rho= 0.9993531

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.00	0.000	-0.008	-0.008	192	0.000	192					=
0.05	0.050										-14
0.20	0.200	0.172	-0.028	933	0.0 %	933					-1
2.50	2.500	2.483	-0.017	10400	0.0 %	10400					0
5.00	5.000	4.987	-0.013	20659	0.0 %	20659					3
7.50	7.500	7.759	0.259	32012	0.0 %	32012					-2
10.0	10.000	9.806	-0.194	40401	0.0 %	40401					

LB135130

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	192	-		0 SFAM01.1	PPB	3/21/2025 14:40	Std
	0.2 S01	933	-		0.2 SFAM01.1	PPB	3/21/2025 14:42	Std
	2.5 S02	10400	-		2.5 SFAM01.1	PPB	3/21/2025 14:44	Std
	5 S03	20659	-		5 SFAM01.1	PPB	3/21/2025 14:46	Std
	7.5 S04	32012	-		7.5 SFAM01.1	PPB	3/21/2025 14:49	Std
	10 S05	40401	-		10 SFAM01.1	PPB	3/21/2025 14:51	Std
ICV056	ICV056	15780	3.7966 -		SFAM01.1	PPB	3/21/2025 14:54	SMPL
ICB056	ICB056	121	-0.0257 -		SFAM01.1	PPB	3/21/2025 14:56	SMPL
CCV047	CCV047	21448	5.1801 -		SFAM01.1	PPB	3/21/2025 14:59	SMPL
CCB047	CCB047	35	-0.0467 -		SFAM01.1	PPB	3/21/2025 15:01	SMPL
PB167265	PBS265	74	-0.0372 -		SFAM01.1	PPB	3/21/2025 15:06	SMPL
Q1546-03	MDL-SOIL	383	0.0382 -		SFAM01.1	PPB	3/21/2025 15:14	SMPL
CCV048	CCV048	21643	5.2277 -		SFAM01.1	PPB	3/21/2025 15:16	SMPL
CCB048	CCB048	251	0.006 -		SFAM01.1	PPB	3/21/2025 15:21	SMPL