

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135135

Analyst By : jignesh

Supervisor Review By : rubina

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/21/2025	15:40
2	CAL2	1	Water	NA	NA	20.2	7.00	03/21/2025	15:41
3	CAL3	1	Water	NA	NA	20.3	10.02	03/21/2025	15:45
4	ICV	1	Water	NA	NA	20.3	7.01	03/21/2025	15:50
5	CCV1	1	Water	NA	NA	20.2	2.01	03/21/2025	16:00
6	Q1625-01	1	Water	NA	NA	20.5	7.29	03/21/2025	16:10
7	Q1625-01DUP	1	Water	NA	NA	20.6	7.30	03/21/2025	16:11
8	CCV2	1	Water	NA	NA	20.3	12.03	03/21/2025	16:15

WORKLIST(Hardcopy Internal Chain)

U3135135

WorkList Name : PH Q1629 WorkList ID : 188459 Department : Wet-Chemistry Date : 03-21-2025 15:23:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1625-01 -- L	FRAC-TANK-A1291	Water	pH	Cool 4 deg C	PSEG03	I11	03/21/2025	9040C

Date/Time 03/21/25 15:30
Raw Sample Received by: JF WDC
Raw Sample Relinquished by: JF WDC

Date/Time 03/21/25 17:00
Raw Sample Received by: JF WDC
Raw Sample Relinquished by: JF WDC