

LB135165


SFAM01.1

INSTRUMENT IN CV

Linear

 μ Abs.:

42852



A= 0.0000e+000

B= 2.3242e-004 *slope*C= -2.7179e-002 *y-intercept*

Rho= 0.9998656

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.032	0.032	254	0.000	254					
0.05	0.050										=
0.20	0.200	0.183	-0.017	905	0.0 %	905					-9
2.50	2.500	2.417	-0.083	10516	0.0 %	10516					-3
5.00	5.000	5.061	0.061	21892	0.0 %	21892					1
7.50	7.500	7.574	0.074	32705	0.0 %	32705					41
10.0	10.000	9.933	-0.067	42852	0.0 %	42852					-1

LB135165

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	254	-		0 SFAM01.1	PPB	3/25/2025 9:36	Std
	0.2 S01	905	-		0.2 SFAM01.1	PPB	3/25/2025 9:38	Std
	2.5 S02	10516	-		2.5 SFAM01.1	PPB	3/25/2025 9:40	Std
	5 S03	21892	-		5 SFAM01.1	PPB	3/25/2025 9:43	Std
	7.5 S04	32705	-		7.5 SFAM01.1	PPB	3/25/2025 9:45	Std
	10 S05	42852	-		10 SFAM01.1	PPB	3/25/2025 9:47	Std
ICV059	ICV059	16231	3.7453 -		SFAM01.1	PPB	3/25/2025 9:51	SMPL
ICB059	ICB059	49	-0.0158 -		SFAM01.1	PPB	3/25/2025 9:53	SMPL
CCV055	CCV055	21849	5.0511 -		SFAM01.1	PPB	3/25/2025 9:55	SMPL
CCB055	CCB055	34	-0.0193 -		SFAM01.1	PPB	3/25/2025 9:58	SMPL
PB167304BL	PBS304	60	-0.0132 -		SFAM01.1	PPB	3/25/2025 10:00	SMPL
Q1546-04	MDL-SOIL	272	0.036 -		SFAM01.1	PPB	3/25/2025 10:05	SMPL
CCV056	CCV056	21775	5.0339 -		SFAM01.1	PPB	3/25/2025 10:07	SMPL
CCB056	CCB056	-27	-0.0335 -		SFAM01.1	PPB	3/25/2025 10:09	SMPL