

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135167

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/25/2025	10:30
2	CAL2	1	Water	NA	NA	20.2	7.00	03/25/2025	10:31
3	CAL3	1	Water	NA	NA	20.3	10.03	03/24/2025	10:35
4	ICV	1	Water	NA	NA	20.2	7.01	03/25/2025	10:40
5	CCV1	1	Water	NA	NA	20.2	2.01	03/25/2025	10:45
6	Q1632-01	1	Solid	20.02	20	23.8	10.61	03/25/2025	11:00
7	Q1632-01DUP	1	Solid	20.03	20	23.9	10.62	03/25/2025	11:02
8	Q1633-01	1	Solid	20.02	20	20.9	6.62	03/25/2025	11:11
9	CCV2	1	Water	NA	NA	20.3	12.02	03/25/2025	11:15

WORKLIST(Hardcopy Internal Chain)

UR 195167

WorkList Name : PH Q1625 S WorkList ID : 188521 Department : Wet-Chemistry Date : 03-25-2025 10:06:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1632-01	PIER-1-2	Solid	pH	Cool 4 deg C	PSEG03	I41	03/24/2025	9045D
Q1633-01	SLG	Solid	pH	Cool 4 deg C	PSEG03	I31	03/24/2025	9045D

Date/Time 03/25/25 10:20

Raw Sample Received by: SP WPC

Raw Sample Relinquished by: OP SR

Date/Time 03/25/25

Raw Sample Received by: OP SR

Raw Sample Relinquished by: SP WPC

12:30

OP SR

SP WPC