

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135173

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	03/25/2025	13:20
2	Q1625-01		1	15	100.00	>212.0	>212.0	03/25/2025	13:50
3	Q1625-01DUP		1	15	100.00	>212.0	>212.0	03/25/2025	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16135173

WorkList Name : FP-3-24

WorkList ID : 188500

Department : Wet-Chemistry

Date : 03-24-2025 16:26:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1625-01	FRAC-TANK-A1291	Water	Flash Point	Cool 4 deg C	PSEG03	I11	03/21/2025	1010B

Date/Time 03/25/2025 13.10
 Raw Sample Received by: R17(wc)
 Raw Sample Relinquished by: R17(wc)

Date/Time 03/25/2025 15.00
 Raw Sample Received by: R17(wc)
 Raw Sample Relinquished by: R17(wc)