

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135186

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	03/26/2025	08:35
2	Q1643-06		1	14	100.00	>212.0	>212.0	03/26/2025	09:05
3	Q1643-06DUP		1	14	100.00	>212.0	>212.0	03/26/2025	09:35
4	Q1643-09		1	16	100.00	>212.0	>212.0	03/26/2025	10:05
5	Q1644-03		1	13	100.00	>212.0	>212.0	03/26/2025	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06